



Certificate of Sterility and Quality Assurance

PRODUCT	LOT NUMBER	EXPIRY DATE	DATE OF CERTIFICATION
MW168S	18g30	31/07/2020	03/08/2018

Fecal Transwab® with Liquid Cary Blair Medium

Parameter	Specification	Result
Sterilisation Method	Irradiation (Procedure conforms to ISO 11137)	PASSED
pH at 25°C	8.0 - 8.5	PASSED
Non-viable count	No more than 2 microorganisms in 10 high power (1000X) fields	PASSED
Performance Testing	Recovery within specification at 25°C for <i>Pseudomonas aeruginosa</i> NCTC 10332 (24 hours) <i>Salmonella typhimurium</i> NCTC 0074 (24 hours) <i>Escherichia coli</i> NCTC 9001 (24 hours)	PASSED

Product has been manufactured in accordance with ISO 9001:2015 and ISO 13485:2003
 Certificate LRQ0939587 A

Product has been manufactured in accordance with the requirements of the Medical Devices Directive 93/42/EEC and the Medical Devices Regulations 2002, UK Statutory Instrument No. 618
 Certificate LRQ0939587 C

Performance testing conducted in accordance with: Internal QC Procedures

For and on behalf of Medical Wire & Equipment

Quality Assurance



QCC1 REVISION 1
ISSUE DATE DEC 16

Medical Wire & Equipment Co (Bath) Ltd

Corsham, Wiltshire, England, SN13 9RT

Tel: +44 1225 810 361

Fax: +44 1225 810 153

info@mwe.co.uk

www.mwe.co.uk



LRQA Notified Body Number
0088