



Certificate of Sterility and Quality Assurance

PRODUCT	LOT NUMBER	EXPIRY DATE	DATE OF CERTIFICATION
MW170	18F21	30/06/2020	25/06/2018

Transwab® with Amies Medium Plain

Parameter	Specification	Result
Sterilisation Method	Irradiation (Procedure conforms to ISO 11137)	PASSED
pH at 25°C	7.3 +/- 0.3	PASSED
Non-viable count	No more than 2 microorganisms in 10 high power (1000X) fields	PASSED
Performance Testing	Recovery within specification at 4°C and 25°C tested with a selection of organisms from the following panel, in accordance with CLSI M40-A2 <i>Pseudomonas aeruginosa</i> ATCC®BAA-427 (48 hours) <i>Streptococcus pyogenes</i> ATCC®19615 (48 hours) <i>Haemophilus influenzae</i> ATCC® 10211 (48 hours) <i>Streptococcus pneumoniae</i> ATCC® 6305 (48 hours) <i>Bacteroides fragilis</i> ATCC® 25285 (48 hours) <i>Peptostreptococcus anaerobius</i> ATCC®27337 (48 hours) <i>Fusobacterium nucleatum</i> ATCC®25586 (48 hours) <i>Prevotella melaninogenica</i> ATCC®25845 (48 hours) <i>Propionibacterium acnes</i> ATCC®6915 (48 hours) <i>Neisseria gonorrhoeae</i> ATCC® 43069 (24 hours)	PASSED

Product has been manufactured in accordance with ISO 9001:2015 and ISO 13485:2003
 Certificate LRQ0939587 A

Product has been manufactured in accordance with the requirements of the Medical Devices Directive 93/42/EEC and the Medical Devices Regulations 2002, UK Statutory Instrument No. 618
 Certificate LRQ0939587 C

Validated performance testing conducted in accordance with: *CLSI Quality Control of Microbiological Transport Systems; Approved Standard- Second Edition CLSI document M40-A2. Wayne, PA: Clinical and Laboratory Standards Institute; 2014*

For and on behalf of Medical Wire & Equipment

A. Nash

Quality Assurance



QCC1 REVISION 1
ISSUE DATE JAN 17

Medical Wire & Equipment Co (Bath) Ltd

Corsham, Wiltshire, England, SN13 9RT

Tel: +44 1225 810 361

Fax: +44 1225 810 153

info@mwe.co.uk

www.mwe.co.uk



LRQA Notified Body Number
0088