



# **Certificate of Sterility and Quality Assurance**

| PRODUCT | LOT NUMBER | EXPIRY DATE | DATE OF CERTIFICATION |
|---------|------------|-------------|-----------------------|
| MW911S  | 18D30      | 30/04/2019  | 03/05/2018            |

*Sigma-VCM™ with Liquid Universal Transportation Medium - Mini Tip Swab*

| Parameter            | Specification   | Result |
|----------------------|---|--------|
| Sterilisation Method | Irradiation<br>(Procedure conforms to ISO 11137)  | PASSED |
| pH at 25°C           | 7.3 +/- 0.3   | PASSED |
| Inhibition Testing   | Inhibition of growth within specification of following organisms:<br><br><i>Escherichia coli</i> NCTC 9001<br><i>Candida albicans</i> NCTC 3179<br><i>Staphylococcus aureus</i> NCTC 8532 | PASSED |

Product has been manufactured in accordance with ISO 9001:2015 and ISO 13485:2003  
*Certificate LRQ0939587 A*

Product has been manufactured in accordance with the requirements of the Medical Devices Directive 93/42/EEC and the Medical Devices Regulations 2002, UK Statutory Instrument No. 618  
*Certificate LRQ0939587 C*

Performance testing conducted in accordance with: Internal Procedures

For and on behalf of Medical Wire & Equipment

Anash

## Quality Assurance



QCC1 REVISION 1  
ISSUE DATE DEC. 16

Medical Wire & Equipment Co (Bath) Ltd

Corsham Wiltshire England SN13 9RT

Tel: +44 1225 810 361

info@mwe.co.uk

Fax: +44 1225 810 153

www.mwe.co.uk



LRQA Notified Body Number  
0088