



## Certificate of Sterility and Quality Assurance

PRODUCT	LOT NUMBER	EXPIRY DATE	DATE OF CERTIFICATION
MWVIM	sa35	01/12/2019	23/01/2018

*Viabank*

Parameter	Specification	Result
Sterility Test	Broth from vials tested by culture methods	<b>PASSED</b>
Visual inspection	Bead numbers checked in 1% of vials	<b>PASSED</b>
Performance testing	Recovery after freezing for: <i>Escherichia coli</i> ATCC ®25922 <i>Legionella pneumophiliia</i> NCTC 11192 <i>Haemophilus influenzae</i> ATCC® 10211 (48 hours) <i>Neisseria gonorrhoeae</i> NCTC 12700	<b>PASSED</b>

Product has been manufactured in accordance with ISO 9001:2015  
Certificate LRQ0939587 A

Testing conducted in accordance with: Internal QC procedures  
All results reported on this certificate were obtained at the time of release for this batch.

For and on behalf of Medical Wire & Equipment

A handwritten signature in black ink, appearing to read 'A. Nash'.

Quality Assurance



QCC1 REVISION 1  
ISSUE DATE JAN 17

**Medical Wire & Equipment Co (Bath) Ltd**

Corsham, Wiltshire, England, SN13 9RT

Tel: +44 1225 810 361

Fax: +44 1225 810 153

info@mwe.co.uk

www.mwe.co.uk



LRQA Notified Body Number  
0088