



CERTIFICATE OF ANALYSIS

PART NO.: C14292
LOT NO.: 8404730
DATE OF MFR.: 9/25/2018-10/1/2018
PRODUCT: Sterile and Certified Free 38.5mL Open Top Ultra-clear Round Bottom Tube, 25 x 89 mm, Qty-50

Certified Free:

Beckman Coulter, Inc. certifies that the product batch described above is manufactured in accordance with manufacturing guidelines and product specifications. The product meets all manufacturing requirements, and is certified to be RNase, DNase, DNA, and Endotoxin free based on sample results below detectable limits. Testing for PCR* inhibition is also performed as part of the DNA free test method. The product is made with 100% virgin resin material.

Animal Origin Statement:

We confirm the raw material used in this product does not contain any animal derived materials in the manufacture or formulation of this product.

Additionally, during processing and manufacturing process, the product does not come in contact or is exposed to any other form of animal derived materials.

Be aware that Beckman Coulter does not test for the presence of these substances and that this statement is based upon documentation from our raw material suppliers.

Product Sterility:

When applicable, the validated sterilization process certifies product labeled as sterile to have a sterility assurance level of 10^{-6} . Product sterility is assured for 3 years from the date of manufacture if package is unopened, undamaged, and stored dry at temperatures between 15° to 25°C (59°-77°F).

Sterilization NO.: 2315976
Sterilization Date: 11/5/2018

Certified by: 
Quality Assurance

Date: 1/11/2019

*The PCR process is covered by patents owned by Roche Molecular Systems, Inc. and F. Hoffman-La Roche, Ltd.