

Corning Incorporated  
Life Sciences

Registered  
ISO 9001:2008

## Product Description

**Catalog Number:** 4141

**Product Description:** DeckWorks™ Pipet Tips, 0.1-10 µL, Graduated, Polypropylene, Natural, Not Sterile, in Tip Station, 2 x 96 Tips/Tip Station Layer, 5 Layers/Tip Station (960 Tips), 5 Packs (Tip Station)/Case (4800 Tips)

**Component Materials:**

- Tip - Polypropylene, virgin, color natural.
- Base - Polypropylene, color natural.
- Tray - Polypropylene, color orange. Heavy metal free color concentrate (meets CONEG requirements).
- Lid - Polypropylene, color natural.

**Product Dimensions:**

Tip: Overall Length 45.3 mm, I.D. (small orifice) 0.38 mm.

Tray: Dual 96 holes in 8 x 12 pattern on 9 mm centers.

Station: Length 168 mm x Width 124 mm x Height 144 mm.

**Sterilization:**

Supplied Not Sterile.

**RNase/DNase Testing:**

This product is tested to confirm that no detectable RNase/DNase contamination is present.

**Pyrogens:**

This product is tested per the criteria established in *Guideline on Validation of the Limulus Amebocyte Lystate Test as an End-Product Endotoxin Test for Human and Parenteral Drugs, Biological Products and Medical Devices* (FDA, December, 1987). Results less than 0.1 EU/mL.

**PCR Certification:**

This product is tested to confirm that no detectable PCR inhibitors are present.

**Autoclaving:**

DeckWorks™ tips, decks, hinged racks, station layer trays, station covers and station bases may be autoclaved using a standard cycle of 121°C and 15 psi for 15 minutes. Place tips into deck/rack or into station trays before autoclaving.

**Lot Number Designation:**

Each production lot of this product is uniquely identified with an 8 digit lot number: First 3 digits - Julian date; Next 2 digits - last 2 digits of year of manufacture; Last 3 digits - Batch identification.

Rev No: 1