


JLA-8.1000 and JLA-9.1000 Rotor Labware Assembly Insert

Polypropylene Bottle Assembly C31597
Polycarbonate Bottle Assembly C31600

Figure 1 Beckman Coulter JLA-8.1000 and JLA-9.1000 Labware Assembly



Please Read Before Use

Contents

This package contains labware containing polypropylene bottle assemblies (C31597) or polycarbonate bottle assemblies (C31600). Assembly contents are listed below.

Polypropylene bottle assembly with polyphenylene sulfide (PPS) cap/closure and Autovent plug assemblies, set of two (C31597) includes:
1000-mL polypropylene bottles, set of two
Cap/closure (PPS) and Autovent plug (PEI) ^a assemblies, set of two
Two small Autovent O-rings
Two large plug O-rings

a. NOTE: Autovent plug assemblies are shipped with one of each O-ring installed, in addition to included O-rings.

Polycarbonate bottle assembly with polyphenylene sulfide (PPS) cap/closure and Autovent plug (PEI), set of two (C31600) includes:
1000-mL polycarbonate bottles, set of two
Cap/closure (PPS) and Autovent plug (PEI) ^a assemblies, set of two
Two small Autovent O-rings
Two large plug O-rings

a. NOTE: Autovent plug assemblies are shipped with one of each O-ring installed, in addition to included O-rings.

Inspection and Usage

The bottles and associated labware contained in this package are designed for use only in the Beckman Coulter JLA-8.1000 and JLA-9.1000 rotors. This labware is different from labware previously manufactured by Beckman Coulter, and requires special inspection and handling procedures. Please read and follow the instructions below. For additional information on the safe use of the rotors, read *J-Lite JLA-8.1000 and JLA-9-1000 Fixed Angle Rotor Assemblies Instructions For Use* (J-TB-073).

- Before each use, inspect each bottle, cap/closure, and plug for cracks or damage. *Do not use damaged components.*
- Do not fill above the maximum fill line marked on the bottle. Minimum fill volume for these bottles is 500 mL.

Below are rated operating conditions for PP and PC bottle assemblies to ensure safe operation.

	JLA-8.1000 Rotor	JLA-9.1000 Rotor
Maximum Speed	8000 rpm	9000 rpm
Maximum Temperature	38°C	38°C
Minimum Temperature	4°C	4°C
Cap Closure Torque (Minimum)	60 in-lbs	60 in-lbs
Minimum Fill Level	500 mL	500 mL
Maximum Fill Level	1000 mL	1000 mL
Maximum filled bottle assembly weight (including bottle, plug, cap/closure, and sample)	1500 Grams	1500 Grams

- *O-ring integrity is key to successful function of the plug.* Before each use, inspect all O-rings to make sure that they are in good condition, dry, and unlubricated. Replace them if they are damaged, torn, or dirty. Extra O-rings of each type are included in this kit, and additional O-rings can be ordered from Beckman Coulter.
- The useful service life of plastic bottles, plugs, and caps has been tested to 100 hours, but will largely depend on handling and usage. Usage includes but is not limited to: Run, duration, speed settings, temperature settings, cleaning and sterilization methods, and chemicals used. Bottle assembly components that show signs of visible wear, damage, stress, or deformation should be replaced. *Do not use damaged components.*
- The manufacturing process can cause small cosmetic flaws on polycarbonate and polypropylene bottles. These flaws, which may appear as irregular or scraped surface areas, are cosmetic only. They do not affect the function nor shorten the life of the bottle.

Replacement Labware

Bottles, 1000 mL polypropylene, set of two	A98814
Bottles, 1000 mL polypropylene, set of twelve	B29048
Bottles, 1000 mL polycarbonate, set of two	366751
Polyphenylene sulfide (PPS) cap/closure replacement assembly, set of six	C31603
Autovent plug (PEI), set of two, includes: Two Autovent plug assemblies ^a Two large plug O-rings Two small Autovent plug O-rings	C31604
Autovent plug replacement O-rings, set of six, includes: Six large plug O-rings Six small Autovent plug O-rings	B10320
Cap/Closure and Plug assemblies, set of two, includes: Two Autovent plug assemblies (PEI) ^a Two Cap/Closure assemblies (PPS) Two large Plug O-Rings Two small Autovent O-Rings	C31602

Tool kit, includes: One bottle grip (366763) One wrench (367309)	366770
1000 mL Bottle Torque Took Kit includes: 1000 mL Torque Adapter (C57316) 1000 mL Bottle Torque Tool Adapter (C57475)	C57474
Spatula (for removing pellet from bottle)	363646

a. NOTE: Plug assemblies are shipped with one of each O-ring installed in addition to included O-rings.

Copyright

© 2020 Beckman Coulter, Inc.
All rights reserved

Beckman Coulter, the stylized logo, and the Beckman Coulter product and service marks mentioned herein are trademarks or registered trademarks of Beckman Coulter, Inc. in the United States and other countries.

All other trademarks, service marks, products, or services are trademarks or registered trademarks of their respective holders.

Contact Us

If you have any questions, contact our Customer Support Center.

- Worldwide, find us via our website at www.beckman.com/support/technical
- In the USA and Canada, call us at 1-800-369-0333.
- Outside of the USA and Canada, contact your local Beckman Coulter Representative.



Beckman Coulter, Inc.
250 S. Kraemer Blvd.
Brea, CA 92821 U.S.A.



Beckman Coulter Eurocenter S.A.
22, rue Juste-Olivier
Case Postale 1044
CH - 1260 Nyon 1, Switzerland
Tel: +41 (0) 22 365 36 11

Glossary of Symbols is available at beckman.com/techdocs (PN C24689).

Original Instructions