

## Certificate of Inspection

**Date** 

**Catalogue Number** 

CC7672-4XXX, CC7682-4XXX

**Lot Number** 

**Product Description** 

CytoOne Flask

STARLAB certifies that the product described above has been manufactured and inspected in accordance with applicable STARLAB International Standard Procedures including following specifications.

TEST/ PROCEDURE	RESULT	SPECIFICATION
Materials	pass	polystyrene
Sterilization	pass	Irradiation date:
Cell attachment and growth	pass	Cell line: TIG-3
Endotoxin	pass	0.1 EU/mL
DNase	pass	$< 1 \text{ X } 10^{-7} \text{ Kunitz units / } \mu\text{I}$
RNase	pass	$< 4.6 \times 10^{-9} \text{ Kunitz units / } \mu\text{l}$
DNA	pass	< = 30 pg
Surface treatment	pass	
Visual attribute	pass	
Packaging	pass	

Approved by

Dr. Lennart Walter

Senior Product Manager