

# CERTIFICATE OF ASSURANCE



This certificate is your assurance that samples of  
TipOne®, TipOne® Repeat, TubeOne® and PlateOne®  
products are manufactured in a regularly controlled environment  
certified free of RNase, DNase, human DNA and Pyrogens.

## *Limits of detection:*

RNase	< 1 x 10 <sup>-9</sup> Kunitz Units
DNase	< 1 x 10 <sup>-6</sup> Kunitz Units
Human DNA	< 2 x 10 <sup>-9</sup> mg
Endotoxins	< 0.05 EU/ml

Sterile products have been sterilised by e-beam irradiation in accordance with  
DIN EN ISO 11137-2:2007 with min. dose 15kGy (1.5 Mrads).

Sterility: SAL 10<sup>-6</sup>.

[www.starlabgroup.com](http://www.starlabgroup.com)

