

CERTIFICATE OF ASSURANCE



This certificate is your assurance that samples of
TipOne®, TipOne® Repeat, TubeOne® and PlateOne®
products are manufactured in a regularly controlled environment
certified free of RNase, DNase, human DNA and Pyrogens.

Limits of detection:

RNase	< 1 x 10 ⁻⁹ Kunitz Units
DNase	< 1 x 10 ⁻⁶ Kunitz Units
Human DNA	< 2 x 10 ⁻⁹ mg
Endotoxins	< 0.05 EU/ml

Sterile products have been sterilised by e-beam irradiation in accordance with
DIN EN ISO 11137-2:2007 with min. dose 15kGy (1.5 Mrads).

Sterility: SAL 10⁻⁶.

www.starlabgroup.com

