

**Corning Incorporated  
Life Sciences**

Registered  
ISO 9001:2008

## Product Description

**Catalog Number:** 3013

**Product Description:** Corning® Micro Spatula Style I, v scoop/round end

**Component Materials:**

Spatula - Anti-static polystyrene and color concentrate blend, meets *USP, Class VI* requirements for plastic containers and closures.

**Product Dimensions:**

Overall Length - 5.9 in. Maximum V Scoop Width - 0.13 in.

**Sterilization:**

This lot has been irradiated and dosimetrically released based on ANSI/AAMI/ISO 11137 (TIR 33) *Sterilization of healthcare products-Requirements for validation and routine control-Radiation sterilization.*

Sterility Assurance Level: SAL 10<sup>-3</sup>

**RNase/DNase Testing:**

This product has been tested and is free of any detectable RNase/DNase contamination.

**Pyrogens:**

The product has been tested and has met the criteria established in ANSI/AAMI ST 72:2002/(R) 2010: *Bacterial Endotoxins - Test methodologies, routine testing, and alternative to batch testing.* Results: less than 0.1 EU/mL

**Performance Testing:**

Each manufacturing lot is sampled and tested in accordance with Standard Operating Procedures.

Visual Attributes:

Visual examination of the product.

Packaging:

Inspection for seal and barrier integrity, accurate labeling, and correct product configuration.

**Lot Number Designation:**

8 Digit Lot Number: First 3 digits - Julian date, start of manufacturing; Next 2 digits - Year of manufacture; Last 3 digits - Batch identification. The entire lot is sterilized as one unit.

Rev No: 5