

**Corning Incorporated
Life Sciences**

Registered
ISO 9001:2008

Product Description

Catalog Number: 3207

Product Description: Costar® 1.7 mL Prelubricated Microcentrifuge Tube

Component Materials:

Tube - Polypropylene, meets *USP, Class VI* requirements for plastic containers and closures.

Product Dimensions:

Length	-	1.535"	I.D. @ top of tube	-	.362"
Tolerances	-	+/- .050"	O.D. @ cap	-	.500"

Special Feature:

The tube is autoclavable.

RNase/DNase Testing:

This product has been tested and is free of any detectable RNase/DNase contamination.

Pyrogens:

The product has been tested and has met the criteria established in the Guidelines on Validation of the Limulus Amebocyte Lysate Test as an End-Product Endotoxin Test for Human and *Animal* Parenteral Drugs, Biological Products and Medical Devices (FDA, December, 1987).

Performance Testing:

Each manufacturing lot is sampled and tested in accordance with Standard Operating Procedures.

Visual Attributes: Visual examination of the product.

Packaging: Accurate labeling, and correct product configuration.

Lot Number Designation:

8 Digit Lot Number: First 3 digits - Julian date, start of manufacturing; Next 2 digits - Year of manufacture; Last 3 digits - Batch identification.

Rev No: 5