

**Corning Incorporated  
Life Sciences**

Registered  
ISO 9001:2008

## Product Description

**Catalog Number:** 11902

**Product Description:** Corning® CellSTACK® Filling Accessory, with Female MPC

**Component Materials:**

Cap	- Virgin High Density Polyethylene, meets <i>USP, Class VI</i> requirements for plastic containers and closures, Heavy metal free (meets CONEG req.) color concentrate.
Tubing	- Thermoplastic Elastomer, meets <i>USP, Class VI</i> requirements for plastic containers and closures.
MPC Connector/Plug	- Polycarbonate, meets <i>USP, Class VI</i> requirements for plastic containers and closures.

**Sterilization:**

This lot has been irradiated and dosimetrically released based on ANSI/AAMI/ISO 11137 (AAMI TIR33) *Sterilization of healthcare products-Requirements for validation and routine control-Radiation sterilization*.

Sterility Assurance Level: SAL 10<sup>-6</sup>

**Bovine Spongiform Encephalopathy and Transmissible Spongiform Encephalopathy:**

This product is manufactured with animal free materials.

**Performance Testing:**

Each manufacturing lot is sampled and tested in accordance with Standard Operating Procedures.

Integrity Test: Pull force test of the Accessory to 10lbs.

Visual Attributes: Visual examination of the product.

Packaging: Inspection for seal and barrier integrity, accurate labeling, and correct product configuration.

**Lot Number Designation:**

8 Digit Lot Number: First 3 digits - Julian date, start of manufacturing; Next 2 digits - Year of manufacture; Last 3 digits - Batch identification.

Rev No: 1