

**Corning Incorporated
Life Sciences**

Registered
ISO 9001:2008

Product Description

Catalog Number: 3362

Product Description: Costar ® 96-well, tissue culture treated, white opaque plate, no lid

Component Materials:

Plate - Virgin Polystyrene, meets *USP, Class VI* requirements for plastic containers and closures. White concentrate

Product Dimensions:

Length of Plate	-	5.030 in.	Diameter of Well @ top	-	.270 in.
Width of Plate	-	3.365 in.	Diameter of Well @ bottom	-	.250 in.
Depth of Well	-	.420 in.	Height	-	0.560 in.
Tolerances of Dimensions	-	+/- .010 in.	Volume per Well	-	.36 mL
Overall Flatness	-	0.0299 in.			

Sterilization:

This lot has been irradiated and dosimetrically released based on ANSI/AAMI/ISO 11137 *Sterilization of healthcare products-Requirements for validation and routine control-Radiation sterilization*.

Sterility Assurance Level: SAL 10⁻³

Surface Characterization:

Surface is characterized to be hydrophilic and negatively charged, composed of 9-17% oxygen atoms. This surface composition has been optimized for cell attachment and growth.

Cell Attachment and Growth Characteristics:

The product has been tested for the attribute of cell attachment and growth utilizing an attachment-dependent mammalian cell line in a serum supplemented media.

Optical Characteristics:

The product is made of opaque white polystyrene walls to minimize well to well crosstalk and background fluorescence and /or luminescence.

Performance Testing:

Each manufacturing lot is sampled and tested in accordance with Standard Operating Procedures.

Visual Attributes:

Visual examination of the product.

Packaging:

Inspection for seal and barrier integrity, accurate labeling and correct product configuration.

Opacity:

Visual, using a fluorescent compound.

Cell Culture Treatment:

Wettability test using water to insure the presence of a hydrophilic surface.

Lot Number Designation:

8 Digit Lot Number: First 3 digits - Julian Date, start of manufacturing; Next 2 digits - Year of manufacture; Last 3 digits - Batch identification.

Rev No: 8