

**Corning Incorporated
Life Sciences**

Registered
ISO 9001:2008

Product Description

Catalog Number: 3747

Product Description: Thermowell® Gold 0.2ml PCR tubes, domed cap, clear

Component Materials:

Tube - Polypropylene

Product Dimensions:

Overall tube length w/out cap	-	.817"	I.D. opening	-	.212"
Tolerances	-	+/- .005"			

RNase/DNase Testing:

This product has been tested and is free of any detectable RNase/DNase contamination.

Pyrogens:

The product has been tested and has met the criteria established in the *Guideline on Validation of the Limulus Amebocyte Lysate Test as an End-Product Endotoxin Test for Human and Animal Parenteral Drugs, Biological Products and Medical Devices* (FDA, December, 1987).

Performance Testing:

Each manufacturing lot is sampled and tested in accordance with Standard Operating Procedures.

Visual Attributes: Visual examination of the product.

Packaging: Inspection for accurate labeling, and correct product configuration.

Lot Number Designation:

8 Digit Lot Number: First 3 digits - Julian date, start of manufacturing; Next 2 digits - Year of manufacture; Last 3 digits - Batch identification.

Rev No: 2