

**Corning Incorporated  
Life Sciences**

Registered  
ISO 9001:2008

## Product Description

**Catalog Number:** 3959

**Product Description:** Costar® 96-well Assay Block, 1 mL, u-bottom, non-sterile

**Component Materials:**

Assay Block - Virgin Polypropylene, meets *USP, Class VI* requirements for plastic containers and closures.

**Product Dimensions:**

Block height	-	1.660 in.	Height of well	-	1.571 in.
Block length @ bottom	-	5.020 in.	Diameter @ top of well	-	.270 in.
Block width @ bottom	-	3.345 in.	Distance from well center to center	-	.354 in.

Tolerances of Dimensions - +/- .010 in.

**Surface Characterization:**

Surface is characterized to be hydrophobic and non-charged.

**RNase/DNase Testing:**

This product has been tested and is free of any detectable RNase/DNase contamination.

**Pyrogens:**

The product has been tested and has met the criteria established in the *Guideline on Validation of the Limulus Amebocyte Lysate Test as an End-Product Endotoxin Test for Human and Animal Parenteral Drugs, Biological Products and Medical Devices (FDA, December, 1987)*.

**Bovine Spongiform Encephalopathy and Transmissible Spongiform Encephalopathy:**

This product complies with the latest revision of EMEA/410/01 "Note for Guidance on minimizing the risk of transmitting animal spongiform encephalopathy agents via human veterinary medicine products" by virtue of all bovine derived material having been processed per specific conditions of section 6.4 of EMEA/410/01.

**Performance Testing:**

Each manufacturing lot is sampled and tested in accordance with Standard Operating Procedures.

Visual Attributes:

Visual examination of the product.

Packaging:

Inspection for accurate labeling and correct product configuration.

**Lot Number Designation:**

8 Digit Lot Number: First 3 digits - Julian date, start of manufacturing; Next 2 digits - Year of manufacture; Last 3 digits - Batch identification.

Rev No: 9