

**Corning Incorporated
Life Sciences**

**Registered
ISO 9001:2008**

Product Description

Catalog Number: 4413

Product Description: Costar ® 1.2 mL Cluster Tube Strip, in rack, sterile

Component Materials:

Tube Strip	-	Polypropylene
Rack	-	Polypropylene

Product Dimensions:

Overall strip length	-	2.820 in.	I.D. @ top	-	0.275 in.
Overall tube length	-	1.830 in.	O.D. @ top	-	0.340 in.
Tolerances	-	+/- .050 in.			

Sterilization:

The lot has been irradiated and dosimetrically released based on ANSI/AAMI/ISO 11137 *Sterilization of health care products-Requirements for validation and routine control-Radiation sterilization*.

Sterility Assurance Level: SAL 10⁻³

RNase/DNase Testing:

This lot has been tested and is free of any detectable RNase/DNase contamination.

Performance Testing:

Each manufacturing lot is sampled and tested in accordance with Standard Operating Procedures.

Visual Attributes: Visual examination of the product.

Packaging: Inspection for seal and barrier integrity, accurate labeling, and correct product configuration.

Lot Number Designation:

8 Digit Lot Number: First 3 digits - Julian date, start of manufacturing; Next 2 digits - Year of manufacture; Last 3 digits - Batch identification.

Rev No: 5