Corning Incorporated Life Sciences

Registered ISO 9001:2008

Product Description



Catalog Number: 430909

Product Description: Corning® 1.5 mL microcentrifuge tube, attached screw cap with o-ring

Component Materials:

Vial Virgin Polypropylene

Cap & Attachment Virgin Polypropylene. Heavy metal free (meets CONEG reg.) color concentrate.

Loop

Silicone, color black O-ring

Product Dimensions:

Length of Tube 47.0 mm Outer Diameter of cap 12.7 mm Length of tube w/cap 48.6 mm Outer Diameter of tube 10.7 mm

Sterilization:

The product has been irradiated and dosimetrically released based on ANSI/AAMI/ISO 11137 Sterilization of healthcare products-Requirements for validation and routine control-Radiation sterilization. Sterility Assurance Level: SAL 10⁻⁵

The product has been tested and has met the criteria established in the Guideline on Validation of the Limulus Amebocyte Lysate Test as an End-Product Endotoxin Test for Human and Animal Parenteral Drugs, Biological Products and Medical Devices (FDA, December, 1987).

Results: less than 0.1 EU/mL

Bovine Spongiform Encephalopathy and Transmissible Spongiform Encephalopathy: This product complies with the latest revision of EMEA/401/01 "Note for Guidance on minimizing the risk of transmitting animal spongiform encephalopathy agents via human veterinary medicine products" by virtue of all bovine derived material having been processed per specific conditions of section 6.4 of EMEA/401/01.

Performance Testing:

Each manufacturing lot is sampled and tested in accordance with Standard Operating Procedures. Inspection for accurate labeling and correct product configuration. Packaging:

Lot Number Designation:

8 Digit Lot Number: First 3 digits - Julian date, start of manufacturing; Next 2 digits - Year of manufacture; Last 3 digits - Run number for that year.

Rev No: 3