



MANUFACTURED IN THE UK

Sterilin Ltd.

Angel Lane Aberbargoed Bargoed
Caerphilly CF81 9FW United Kingdom

Customer services:

Tel +44 (0)844 844 3737 Fax +44 (0)844 844 2373

Head office:

Tel +44 (0)1443 830 830 Fax +44 (0)1433 821 545

Registered No. 6512031

BS EN ISO 9001 registered company

TECHNICAL DATA SHEET

Product Group

Polystyrene Petri Dishes

Material Composition

Lid & Base

General Purpose Virgin Polystyrene

ALL THE ABOVE MATERIALS ARE FDA APPROVED.

Manufacture

- 1 All Polystyrene Petri Dishes are manufactured in compliance with product and quality requirements as documented in the BS EN ISO 9001:2008 Quality System.
- 2 Verification of product clean room conditions in accordance with BS EN ISO 14644 'Environmental Cleanliness in enclosed spaces'
- 3 Petri Dishes are produced and tested in accordance with the BS611 PART 2 standard, which includes stringent dimensional controls.
- 4 Sterilin Petri Dishes are CE marked conforming to the In-vitro diagnostic directive (98/79/EC)

Sterilisation

- 1 Also available in a sterile format, sterilised by gamma irradiation (dosage 10 – 30 kGy)
2. Sterility is obtained through irradiation according to ISO 11137 (sterilisation of healthcare products - requirements for validation and routine control - radiation sterilisation) and EN552 (sterilisation of Medical Devices – validation and routine control of sterilisation by irradiation).



IWAKI

