

L A N D A N G E R	FICHE TECHNIQUE PRODUIT <i>PRODUCT TECHNICAL SHEET</i>	FAQ 026 INFO Indice : 2 Date : 01/06/2018
-------------------	---	---

Éditée le / *Publishing date*

2022-04-28

Page

1

REFERENCE	DESIGNATIONS - <i>DESCRIPTIONS</i>	CODE EAN
A902010	CANULE ASPIRATION ANDREW-PYNCHON L:230MM D:6X3.4MM	3661219195863
	ANDREW-PYNCHON SUCTION CANNULA L:230MM D:6X3.4MM	

Catalogues / *catalogs*

Landanger 2015

Chapitre-page

15-026

DESCRIPTIF DU DISPOSITIF/ *PRODUCT DESCRIPTION*

XV - Divers
 XV - Miscellaneous
 XV - Diverses

L A N D A N G E R

Canules aspiratrices - Suction tubes - Cánulas aspirantes

Copyright © LANDANGER - 2009 - 2015

SANTY

Ø 9 x 6 mm

A90015
26 cm - 10 1/4"

YANKAUER

Ø 10 x 6 mm

A90090
28 cm - 11"

NUBOER

Ø 6 mm

A902000
21 cm - 8 1/4"

ANDREW PYNCHON

Ø 6 x 3,4 mm

A902010
23 cm - 9"

AUBERT (FINSTERER)

Ø 15 x 10 mm

A90300
30 cm - 12"

BABY POOLE

Ø 6 mm

A903100
20 cm - 8"

POOLE

Ø 8 mm

A903110
22 cm - 8 3/4"

POOLE

Ø 10 mm

A903120
22 cm - 8 3/4"

XV - 26

 1/2

LANDANGER	FICHE TECHNIQUE PRODUIT <i>PRODUCT TECHNICAL SHEET</i>	FAQ 026 INFO Indice : 2 Date : 01/06/2018
-----------	---	---

Éditée le / *Publishing date*

2022-04-28

Page

2

REFERENCE	DESIGNATIONS - <i>DESCRIPTIONS</i>	CODE EAN
A902010	CANULE ASPIRATION ANDREW-PYNCHON L:230MM D:6X3.4MM	3661219195863
	ANDREW-PYNCHON SUCTION CANNULA L:230MM D:6X3.4MM	 3 661219 195863

INFORMATIONS REGLEMENTAIRES / *REGULATORY INFORMATION*

Classe CE :

Ia

Code GMDN :

57822

CE Class :

GMDN code :

PERFORMANCES REVENDIQUEES / *INTENDED PERFORMANCE*

Dispositif médical de bloc opératoire :

REUTILISABLE/REUSABLE

Operating room medical device :

CARACTERISTIQUES SPECIFIQUE / *SPECIFIC FEATURES*

Type matière :

ACIER INOXYDABLE / STAINLESS STEEL

Material type :

Nuance matière :

X5CrNi18-10(1.4301)

Material grade :

Dimensions :

Voir désignations produits

Landanger 2015

Outer dimensions :

See products description

Chapitre-pages

15-026

Nettoyage/Pré-désinfection : Standard pour instrumentation de bloc opératoire réutilisable.

Cleaning / Pre-disinfection : *Standard conditions for reusable surgical instruments*

Stérilisation :

La stérilisation doit être effectuée selon un procédé de stérilisation à la vapeur conformément à la norme ISO 17665-1 (procédé sous vide fractionné) avec un stérilisateur conformément à la norme EN 285.

Méthode de stérilisation à la vapeur saturée	Température	Durée	Temps de séchage	Pression absolue
Pré-vide	132 °C (270 °F)	4 minutes	30 minutes	2,9 bars, 2,8 atm, 41,6 psi
Pré-vide	134 °C (273 °F)	3 minutes	35 minutes	3,0 bars, 3,0 atm, 44,1 psi
Pré-vide	134 °C (273 °F)	5 minutes	35 minutes	3,0 bars, 3,0 atm, 44,1 psi
Pré-vide	134 °C (273 °F)	18 minutes	35 minutes	3,0 bars, 3,0 atm, 44,1 psi

Sterilization process :

The Sterilization shall be performed according to ISO 17665-1 validated steam sterilisation procedure (fractional vacuum procedure) in an autoclave in compliance with EN 285.

Saturated steam sterilization method	Temperature	Time	Drying time	Absolute pressure
Pre-vacuum	132 °C (270 °F)	4 minutes	30 minutes	2.9 bar, 2.8 atm, 41.6 psi
Pre-vacuum	134 °C (273 °F)	3 minutes	35 minutes	3.0 bar, 3.0 atm, 44.1 psi
Pre-vacuum	134 °C (273 °F)	5 minutes	35 minutes	3.0 bar, 3.0 atm, 44.1 psi
Pre-vacuum	134 °C (273 °F)	18 minutes	35 minutes	3.0 bar, 3.0 atm, 44.1 psi

EMBALLAGE / *PACKAGING*

Conditionnement :

Pochette plastique en polyéthylène (PE).

Packaging :

Polyethylene plastic pouch (PE).