

L A N D A N G E R	FICHE TECHNIQUE PRODUIT PRODUCT TECHNICAL SHEET	FAQ 026 INFO Indice : 2 Date : 01/06/2018
-------------------	--	--

Editée le / Publishing date

2022-04-28

Page

1

REFERENCE	DESIGNATIONS - DESCRIPTIONS	CODE EAN
B40830	PINCE DISSECTION TRES FINE A GRIFFES L:110MM	3661219007470
	DISSECTING FORCEPS VERY FINE WITH TEETH L:110MM	 3 661219 007470

Catalogues / catalogs

Landanger 2015

Chapitre-page

02-036

DESCRIPTIF DU DISPOSITIF/ PRODUCT DESCRIPTION

II - Dissection
 II - Dissection
 II - Dissection

L A N D A N G E R

Pinces à dissection ophtalmologiques - Ophthalmologic dissecting forceps - Pinzas para disección oftalmológicas

Copyright © LANDANGER - 2009 - 2015


I150880
 10,5 cm - 4 1/4"

Sans griffes
Without teeth
Sin dientes


I150881
 10,5 cm - 4 1/4"

Sans griffes
Without teeth
Sin dientes


I150882
 10,5 cm - 4 1/4"

1 x 2 griffes
1 x 2 teeth
1 x 2 dientes


I150885
 10,5 cm - 4 1/4"

1 x 2 griffes
1 x 2 teeth
1 x 2 dientes


I150886
 10,5 cm - 4 1/4"

1 x 2 griffes
1 x 2 teeth
1 x 2 dientes


I150887
 10,5 cm - 4 1/4"

Sans griffes
Mors 8/10
Without teeth
8/10 jaws
Sin dientes
Bocas 8/10


I150890
 10 cm - 4"

Sans griffes
Without teeth
Sin dientes


I150895
 10 cm - 4"

Sans griffes
Without teeth
Sin dientes


B40840
 11 cm - 4 1/4"

1 x 2 griffes
1 x 2 teeth
1 x 2 dientes


I15120
 11 cm - 4 1/4"

1 x 2 griffes
1 x 2 teeth
1 x 2 dientes


B40830
 11 cm - 4 1/4"

1 x 2 griffes
1 x 2 teeth
1 x 2 dientes


I15110
 11 cm - 4 1/4"

II - 36

 1/1

LANDANGER	FICHE TECHNIQUE PRODUIT <i>PRODUCT TECHNICAL SHEET</i>	FAQ 026 INFO Indice : 2 Date : 01/06/2018
-----------	---	---

Éditée le / *Publishing date*

2022-04-28

Page

2

REFERENCE	DESIGNATIONS - <i>DESCRIPTIONS</i>	CODE EAN
B40830	PINCE DISSECTION TRES FINE A GRIFFES L:110MM	3661219007470
	DISSECTING FORCEPS VERY FINE WITH TEETH L:110MM	 3 661219 007470

INFORMATIONS REGLEMENTAIRES / *REGULATORY INFORMATION*

Classe CE :
CE Class :

I

Code GMDN :
GMDN code :

62466

PERFORMANCES REVENDIQUEES / *INTENDED PERFORMANCE*

Dispositif médical de bloc opératoire :
Operating room medical device :

REUTILISABLE/REUSABLE

CARACTERISTIQUES SPECIFIQUE / *SPECIFIC FEATURES*

Type matière :
Material type :

ACIER INOXYDABLE / STAINLESS STEEL

Nuance matière :
Material grade :

X20Cr13(1.4021)

Dimensions : Voir désignations produits
Outer dimensions : See products description
Chapitre-pages

Landanger 2015

02-036

Nettoyage/Pré-désinfection : Standard pour instrumentation de bloc opératoire réutilisable.
Cleaning / Pre-disinfection : Standard conditions for reusable surgical instruments

Stérilisation :

La stérilisation doit être effectuée selon un procédé de stérilisation à la vapeur conformément à la norme ISO 17665-1 (procédé sous vide fractionné) avec un stérilisateur conformément à la norme EN 285.

Méthode de stérilisation à la vapeur saturée	Température	Durée	Temps de séchage	Pression absolue
Pré-vide	132 °C (270 °F)	4 minutes	30 minutes	2,9 bars, 2,8 atm, 41,6 psi
Pré-vide	134 °C (273 °F)	3 minutes	35 minutes	3,0 bars, 3,0 atm, 44,1 psi
Pré-vide	134 °C (273 °F)	5 minutes	35 minutes	3,0 bars, 3,0 atm, 44,1 psi
Pré-vide	134 °C (273 °F)	18 minutes	35 minutes	3,0 bars, 3,0 atm, 44,1 psi

Sterilization process :

The Sterilization shall be performed according to ISO 17665-1 validated steam sterilisation procedure (fractional vacuum procedure) in an autoclave in compliance with EN 285.

Saturated steam sterilization method	Temperature	Time	Drying time	Absolute pressure
Pre-vacuum	132 °C (270 °F)	4 minutes	30 minutes	2.9 bar, 2.8 atm, 41.6 psi
Pre-vacuum	134 °C (273 °F)	3 minutes	35 minutes	3.0 bar, 3.0 atm, 44.1 psi
Pre-vacuum	134 °C (273 °F)	5 minutes	35 minutes	3.0 bar, 3.0 atm, 44.1 psi
Pre-vacuum	134 °C (273 °F)	18 minutes	35 minutes	3.0 bar, 3.0 atm, 44.1 psi

EMBALLAGE / *PACKAGING*

Conditionnement : Pochette plastique en polyéthylène (PE).
Packaging : Polyethylene plastic pouch (PE).