

L A N D A N G E R	FICHE TECHNIQUE PRODUIT <i>PRODUCT TECHNICAL SHEET</i>	FAQ 026 INFO Indice : 2 Date : 01/06/2018
-------------------	---	---

Éditée le / *Publishing date*

2022-04-28

Page

1

REFERENCE	DESIGNATIONS - <i>DESCRIPTIONS</i>	CODE EAN
B60720	PASSE FIL O'SHAUGNESSY L:180MM FORT COURBE 90 °	3661219009429
	O'SHAUGNESSY LIGATURE FORCEPS L:180MM STRONG CURVED 90 °	 3 661219 009429

Catalogues / *catalogs*

Landanger 2015

Chapitre-page

04-052

DESCRIPTIF DU DISPOSITIF/ *PRODUCT DESCRIPTION*

IV - Pincas
 IV - Forceps
 IV - Pinzas

L A N D A N G E R

Pincas passe-fils, dissecteurs - Ligature forceps, dissectors - Disecadores

O'SHAUGNESSY		
	45° Fort 45° Strong 45° Fuerte	90° Fort 90° Strong 90° Fuerte
14 cm - 5 1/2 "	B60610	B60710
16 cm - 6 1/4 "	B60615	B60715
18 cm - 7 "	B60620	B60720
20 cm - 8 "	B60630	B60730
22 cm - 8 3/4 "	B60640	B60740
24 cm - 9 1/2 "	B60650	B60750
28 cm - 11 "	B60660	B60752
30 cm - 11 3/4 "	B60670	B60753



B60615
16 cm - 6 1/4 "



B60715
16 cm - 6 1/4 "

Copyright © LANDANGER - 2009 - 2015

IV - 52



LANDANGER	FICHE TECHNIQUE PRODUIT <i>PRODUCT TECHNICAL SHEET</i>	FAQ 026 INFO Indice : 2 Date : 01/06/2018
-----------	---	---

Éditée le / *Publishing date*

2022-04-28

Page

2

REFERENCE	DESIGNATIONS - <i>DESCRIPTIONS</i>	CODE EAN
B60720	PASSE FIL O'SHAUGNESSY L:180MM FORT COURBE 90 °	3661219009429
	O'SHAUGNESSY LIGATURE FORCEPS L:180MM STRONG CURVED 90 °	 3 661219 009429

INFORMATIONS REGLEMENTAIRES / *REGULATORY INFORMATION*

Classe CE :

I

Code GMDN :

62468

CE Class :

GMDN code :

PERFORMANCES REVENDIQUEES / *INTENDED PERFORMANCE*

Dispositif médical de bloc opératoire :

REUTILISABLE/REUSABLE

Operating room medical device :

CARACTERISTIQUES SPECIFIQUE / *SPECIFIC FEATURES*

Type matière :

ACIER INOXYDABLE / STAINLESS STEEL

Material type :

Nuance matière :

X20Cr13(1.4021)

Material grade :

Dimensions :

Voir désignations produits

Landanger 2015

Outer dimensions :

See products description

Chapitre-pages

04-052

Nettoyage/Pré-désinfection : Standard pour instrumentation de bloc opératoire réutilisable.

Cleaning / Pre-disinfection : *Standard conditions for reusable surgical instruments*

Stérilisation :

La stérilisation doit être effectuée selon un procédé de stérilisation à la vapeur conformément à la norme ISO 17665-1 (procédé sous vide fractionné) avec un stérilisateur conformément à la norme EN 285.

Méthode de stérilisation à la vapeur saturée	Température	Durée	Temps de séchage	Pression absolue
Pré-vide	132 °C (270 °F)	4 minutes	30 minutes	2,9 bars, 2,8 atm, 41,6 psi
Pré-vide	134 °C (273 °F)	3 minutes	35 minutes	3,0 bars, 3,0 atm, 44,1 psi
Pré-vide	134 °C (273 °F)	5 minutes	35 minutes	3,0 bars, 3,0 atm, 44,1 psi
Pré-vide	134 °C (273 °F)	18 minutes	35 minutes	3,0 bars, 3,0 atm, 44,1 psi

Sterilization process :

The Sterilization shall be performed according to ISO 17665-1 validated steam sterilisation procedure (fractional vacuum procedure) in an autoclave in compliance with EN 285.

Saturated steam sterilization method	Temperature	Time	Drying time	Absolute pressure
Pre-vacuum	132 °C (270 °F)	4 minutes	30 minutes	2.9 bar, 2.8 atm, 41.6 psi
Pre-vacuum	134 °C (273 °F)	3 minutes	35 minutes	3.0 bar, 3.0 atm, 44.1 psi
Pre-vacuum	134 °C (273 °F)	5 minutes	35 minutes	3.0 bar, 3.0 atm, 44.1 psi
Pre-vacuum	134 °C (273 °F)	18 minutes	35 minutes	3.0 bar, 3.0 atm, 44.1 psi

EMBALLAGE / *PACKAGING*

Conditionnement :

Pochette plastique en polyéthylène (PE).

Packaging :

Polyethylene plastic pouch (PE).