

L A N D A N G E R	FICHE TECHNIQUE PRODUIT <i>PRODUCT TECHNICAL SHEET</i>	FAQ 026 INFO Indice : 2 Date : 01/06/2018
-------------------	---	---

Éditée le / *Publishing date*

2022-04-29

Page

1

REFERENCE	DESIGNATIONS - <i>DESCRIPTIONS</i>	CODE EAN
D035046	MICRO PINCE ALLIGATOR 0,8MM L:80MM, MORS STRIES L:4,5MM	3661219399780
	ALLIGATOR MICRO FORCEPS 0,8MM L:80MM, L:4,5MM SERRATED JAWS	 3 661219 399780

Catalogues / *catalogs*

ORL 2021

Chapitre-page

158

DESCRIPTIF DU DISPOSITIF/ *PRODUCT DESCRIPTION*

L A N D A N G E R

Micro pincers - Micro forceps - Micro pinzas

ALLIGATOR

 D035045 0,8 x 3 mm

 D035046 0,8 x 4,5 mm

 D035045
13 cm - 5"

BELLUCCI

 D036012 1,6 x 8 mm

 D036012
14 cm - 5 1/2"

NOYES

 D032039

 D032039
14 cm - 5 1/2"

158

 1/1

Wt. Catalogue
Wt. Catalogue
Wt. Catalogue

Copyright © LANDANGER - 2018 - 2021

LANDANGER	FICHE TECHNIQUE PRODUIT <i>PRODUCT TECHNICAL SHEET</i>	FAQ 026 INFO Indice : 2 Date : 01/06/2018
-----------	---	---

Éditée le / *Publishing date*

2022-04-29

Page

2

REFERENCE	DESIGNATIONS - <i>DESCRIPTIONS</i>	CODE EAN
D035046	MICRO PINCE ALLIGATOR 0,8MM L:80MM, MORS STRIES L:4,5MM	3661219399780
	ALLIGATOR MICRO FORCEPS 0,8MM L:80MM, L:4,5MM SERRATED JAWS	 3 661219 399780

INFORMATIONS REGLEMENTAIRES / *REGULATORY INFORMATION*

Classe CE :
CE Class :

I

Code GMDN :
GMDN code :

62468

PERFORMANCES REVENDIQUEES / *INTENDED PERFORMANCE*

Dispositif médical de bloc opératoire :
Operating room medical device :

REUTILISABLE/REUSABLE

CARACTERISTIQUES SPECIFIQUE / *SPECIFIC FEATURES*

Type matière :
Material type :

ACIER INOXYDABLE / STAINLESS STEEL

Nuance matière :
Material grade :

X20Cr13(1.4021) + X8CrNiS18-9(1.4305) + X10CrNi18-8(1.4310)

Dimensions :
Outer dimensions :

Voir désignations produits
See products description
Chapitre-pages

ORL 2021

158

Nettoyage/Pré-désinfection :
Cleaning / Pre-disinfection :

Standard pour instrumentation de bloc opératoire réutilisable.
Standard conditions for reusable surgical instruments

Stérilisation :

La stérilisation doit être effectuée selon un procédé de stérilisation à la vapeur conformément à la norme ISO 17665-1 (procédé sous vide fractionné) avec un stérilisateur conformément à la norme EN 285.

Méthode de stérilisation à la vapeur saturée	Température	Durée	Temps de séchage	Pression absolue
Pré-vide	132 °C (270 °F)	4 minutes	30 minutes	2,9 bars, 2,8 atm, 41,6 psi
Pré-vide	134 °C (273 °F)	3 minutes	35 minutes	3,0 bars, 3,0 atm, 44,1 psi
Pré-vide	134 °C (273 °F)	5 minutes	35 minutes	3,0 bars, 3,0 atm, 44,1 psi
Pré-vide	134 °C (273 °F)	18 minutes	35 minutes	3,0 bars, 3,0 atm, 44,1 psi

Sterilization process :

The Sterilization shall be performed according to ISO 17665-1 validated steam sterilisation procedure (fractional vacuum procedure) in an autoclave in compliance with EN 285.

Saturated steam sterilization method	Temperature	Time	Drying time	Absolute pressure
Pre-vacuum	132 °C (270 °F)	4 minutes	30 minutes	2.9 bar, 2.8 atm, 41.6 psi
Pre-vacuum	134 °C (273 °F)	3 minutes	35 minutes	3.0 bar, 3.0 atm, 44.1 psi
Pre-vacuum	134 °C (273 °F)	5 minutes	35 minutes	3.0 bar, 3.0 atm, 44.1 psi
Pre-vacuum	134 °C (273 °F)	18 minutes	35 minutes	3.0 bar, 3.0 atm, 44.1 psi

EMBALLAGE / *PACKAGING*

Conditionnement :
Packaging :

Pochette plastique en polyéthylène (PE).
Polyethylene plastic pouch (PE).