

LANDANGER	FICHE TECHNIQUE PRODUIT <i>PRODUCT TECHNICAL SHEET</i>	FAQ 026 INFO Indice : 2 Date : 01/06/2018
-----------	---	---

Éditée le / *Publishing date*

2022-04-28

Page

1

REFERENCE	DESIGNATIONS - <i>DESCRIPTIONS</i>	CODE EAN
F251635	VALVE VAGINALE SIMS L:170MM 63X32MM	3661219342236
	SIMS VAGINAL BLADE L:170MM 63X32MM	 3 661219 342236

Catalogues / *catalogs*

Landanger 2015

Chapitre-page

08-045

DESCRIPTIF DU DISPOSITIF/ *PRODUCT DESCRIPTION*

LANDANGER

Valves vaginales - Vaginal blades - Valvas vaginales

SIMS

A l'unité - Per unit - La unidad

	Longueur Length Largo	Largeur Width Anchura	Longueur Length Largo	Largeur Width Anchura
F251631	60 mm	25 mm	60 mm	30 mm
F251632	70 mm	30 mm	70 mm	35 mm
F251633	80 mm	35 mm	80 mm	40 mm

Jeux de 3 - Set of 3 - Juego de 3

 **F251630**

	Longueur Length Largo	Largeur Width Anchura
 F251635	63 mm	32 mm
F251636	76 mm	32 mm
F251637	89 mm	38 mm



F251630
18 cm - 7" ¹⁸



F251635
17 cm - 6 3/4" ¹⁷

VIII - 45

VIII - Vagina bases
VIII - Gynectomy/Procto
VIII - Gynectomy/Procto

Copyright © LANDANGER - 2009 - 2015

4/5

LANDANGER	FICHE TECHNIQUE PRODUIT <i>PRODUCT TECHNICAL SHEET</i>	FAQ 026 INFO Indice : 2 Date : 01/06/2018
-----------	---	---

Éditée le / *Publishing date*

2022-04-28

Page

2

REFERENCE	DESIGNATIONS - <i>DESCRIPTIONS</i>	CODE EAN
F251635	VALVE VAGINALE SIMS L:170MM 63X32MM	3661219342236
	SIMS VAGINAL BLADE L:170MM 63X32MM	 3 661219 342236

INFORMATIONS REGLEMENTAIRES / *REGULATORY INFORMATION*

Classe CE :
CE Class :

I-MDR

Code GMDN :
GMDN code :

45918

PERFORMANCES REVENDIQUEES / *INTENDED PERFORMANCE*

Dispositif médical de bloc opératoire :
Operating room medical device :

REUTILISABLE/REUSABLE

CARACTERISTIQUES SPECIFIQUE / *SPECIFIC FEATURES*

Type matière :
Material type :

ACIER INOXYDABLE / STAINLESS STEEL

Nuance matière :
Material grade :

X2CrNi19-11(1.4306)

Dimensions : Voir désignations produits
Outer dimensions : See products description
Chapitre-pages

Landanger 2015

08-045

Nettoyage/Pré-désinfection : Standard pour instrumentation de bloc opératoire réutilisable.
Cleaning / Pre-disinfection : Standard conditions for reusable surgical instruments

Stérilisation :

La stérilisation doit être effectuée selon un procédé de stérilisation à la vapeur conformément à la norme ISO 17665-1 (procédé sous vide fractionné) avec un stérilisateur conformément à la norme EN 285.

Méthode de stérilisation à la vapeur saturée	Température	Durée	Temps de séchage	Pression absolue
Pré-vide	132 °C (270 °F)	4 minutes	30 minutes	2,9 bars, 2,8 atm, 41,6 psi
Pré-vide	134 °C (273 °F)	3 minutes	35 minutes	3,0 bars, 3,0 atm, 44,1 psi
Pré-vide	134 °C (273 °F)	5 minutes	35 minutes	3,0 bars, 3,0 atm, 44,1 psi
Pré-vide	134 °C (273 °F)	18 minutes	35 minutes	3,0 bars, 3,0 atm, 44,1 psi

Sterilization process :

The Sterilization shall be performed according to ISO 17665-1 validated steam sterilisation procedure (fractional vacuum procedure) in an autoclave in compliance with EN 285.

Saturated steam sterilization method	Temperature	Time	Drying time	Absolute pressure
Pre-vacuum	132 °C (270 °F)	4 minutes	30 minutes	2.9 bar, 2.8 atm, 41.6 psi
Pre-vacuum	134 °C (273 °F)	3 minutes	35 minutes	3.0 bar, 3.0 atm, 44.1 psi
Pre-vacuum	134 °C (273 °F)	5 minutes	35 minutes	3.0 bar, 3.0 atm, 44.1 psi
Pre-vacuum	134 °C (273 °F)	18 minutes	35 minutes	3.0 bar, 3.0 atm, 44.1 psi

EMBALLAGE / *PACKAGING*

Conditionnement : Pochette plastique en polyéthylène (PE).
Packaging : Polyethylene plastic pouch (PE).