

LANDANGER	FICHE TECHNIQUE PRODUIT PRODUCT TECHNICAL SHEET	FAQ 026 INFO Indice : 2 Date : 01/06/2018
-----------	--	---

Éditée le / Publishing date

2022-04-28

Page

1

REFERENCE	DESIGNATIONS - DESCRIPTIONS	CODE EAN
G305100	PORTE AIGUILLE MATHIEU L:140MM MORS TUNGSTENE	3661219343363
	MATHIEU NEEDLE HOLDER L:140MM TUNGSTEN	 3 661219 343363

Catalogues / catalogs

Landanger 2015

Chapitre-page

06-025

DESCRIPTIF DU DISPOSITIF/ PRODUCT DESCRIPTION

LANDANGER

Porte-aiguilles - Needle holders - Porta agujas

MATHIEU



TC Inserts

14 cm - 5 1/2"	G305000	14 cm - 5 1/2"	G305100
17 cm - 6 3/4"	G305010	17 cm - 6 3/4"	G305110
20 cm - 8"	G305015	20 cm - 8"	G305115
24 cm - 9 1/2"	G305020	24 cm - 9 1/2"	G305120



G305010
17 cm - 6 3/4"



G305110
17 cm - 6 3/4"



G305215
20 cm - 8"

Copyright © LANDANGER - 2009 - 2015

VI - Sutures
VI - Sutures
VI - Sutura

VI - 25

LANDANGER	FICHE TECHNIQUE PRODUIT <i>PRODUCT TECHNICAL SHEET</i>	FAQ 026 INFO Indice : 2 Date : 01/06/2018
-----------	---	---

Éditée le / *Publishing date*

2022-04-28

Page

2

REFERENCE	DESIGNATIONS - <i>DESCRIPTIONS</i>	CODE EAN
G305100	PORTE AIGUILLE MATHIEU L:140MM MORS TUNGSTENE	3661219343363
	MATHIEU NEEDLE HOLDER L:140MM TUNGSTEN	 3 661219 343363

INFORMATIONS REGLEMENTAIRES / *REGULATORY INFORMATION*

Classe CE :
CE Class :

I

Code GMDN :
GMDN code :

12726

PERFORMANCES REVENDIQUEES / *INTENDED PERFORMANCE*

Dispositif médical de bloc opératoire :
Operating room medical device :

REUTILISABLE/REUSABLE

CARACTERISTIQUES SPECIFIQUE / *SPECIFIC FEATURES*

Type matière :
Material type :

ACIER INOXYDABLE + TUNGSTENE / STAINLESS STEEL + TUNGSTEN

Nuance matière :
Material grade :

X20Cr13(1.4021) + TUNGSTENE

Dimensions : Voir désignations produits
Outer dimensions : *See products description*
Chapitre-pages

Landanger 2015

06-025

Nettoyage/Pré-désinfection : Standard pour instrumentation de bloc opératoire réutilisable.
Cleaning / Pre-disinfection : *Standard conditions for reusable surgical instruments*

Stérilisation :

La stérilisation doit être effectuée selon un procédé de stérilisation à la vapeur conformément à la norme ISO 17665-1 (procédé sous vide fractionné) avec un stérilisateur conformément à la norme EN 285.

Méthode de stérilisation à la vapeur saturée	Température	Durée	Temps de séchage	Pression absolue
Pré-vide	132 °C (270 °F)	4 minutes	30 minutes	2,9 bars, 2,8 atm, 41,6 psi
Pré-vide	134 °C (273 °F)	3 minutes	35 minutes	3,0 bars, 3,0 atm, 44,1 psi
Pré-vide	134 °C (273 °F)	5 minutes	35 minutes	3,0 bars, 3,0 atm, 44,1 psi
Pré-vide	134 °C (273 °F)	18 minutes	35 minutes	3,0 bars, 3,0 atm, 44,1 psi

Sterilization process :

The Sterilization shall be performed according to ISO 17665-1 validated steam sterilisation procedure (fractional vacuum procedure) in an autoclave in compliance with EN 285.

Saturated steam sterilization method	Temperature	Time	Drying time	Absolute pressure
Pre-vacuum	132 °C (270 °F)	4 minutes	30 minutes	2.9 bar, 2.8 atm, 41.6 psi
Pre-vacuum	134 °C (273 °F)	3 minutes	35 minutes	3.0 bar, 3.0 atm, 44.1 psi
Pre-vacuum	134 °C (273 °F)	5 minutes	35 minutes	3.0 bar, 3.0 atm, 44.1 psi
Pre-vacuum	134 °C (273 °F)	18 minutes	35 minutes	3.0 bar, 3.0 atm, 44.1 psi

EMBALLAGE / *PACKAGING*

Conditionnement :
Packaging :

Pochette plastique en polyéthylène (PE).
Polyethylene plastic pouch (PE).