

L A N D A N G E R	FICHE TECHNIQUE PRODUIT PRODUCT TECHNICAL SHEET	FAQ 026 INFO Indice : 2 Date : 01/06/2018
-------------------	--	--

Editée le / Publishing date

2022-04-29

Page

1

REFERENCE	DESIGNATIONS - <i>DESCRIPTIONS</i>	CODE EAN
I010120	PINCE LARYNGEE MORS CURETTE D:4MM L:220MM DROITE	3661219401988
	MICRO LARYNGEAL FORCEPS ROUND CUPS D:4MM L:220MM STRAIGHT	 3 661219 401988

Catalogues / catalogs

ORL 2021

Chapitre-page

319

DESCRIPTIF DU DISPOSITIF/ *PRODUCT DESCRIPTION*

L A N D A N G E R

Biopsie & corps étrangers - Biopsy & foreign bodies - Biopsia y cuerpos extraños

350 mm - Ø 2.0 mm



3 mm 3 mm 2x5 mm Alligator


I010240


I010250


I010260


I010290

I010290
44 cm - 17 1/4"

L - Ø 4.0 mm



4 mm 4 mm 3x8 mm Alligator

	I010241	I010251	I010261	I010291
200 mm		I010120	I010121	-
220 mm	I010242	I010252	I010262	I010292
250 mm	I010243	I010253	I010263	I010293
300 mm	I010244	I010254	I010264	I010294
350 mm	I010245	I010255	I010265	I010295
400 mm	I010246	I010256	I010266	I010296
450 mm	I010247	I010257	I010267	I010297
500 mm	I010248	I010258	I010268	I010298
550 mm	I010249	I010259	I010269	I010299
600 mm				

I010120
32 cm - 12 3/4"

Courbe droite
Right curve
Curva derecha
4 mm



I010122

Courbe gauche
Left curve
Curva izquierda
4 mm



I010123

Copyright © LANDANGER - 2018 - 2021

2/3

LANDANGER	FICHE TECHNIQUE PRODUIT <i>PRODUCT TECHNICAL SHEET</i>	FAQ 026 INFO Indice : 2 Date : 01/06/2018
-----------	---	---

Éditée le / *Publishing date*

2022-04-29

Page

2

REFERENCE	DESIGNATIONS - <i>DESCRIPTIONS</i>	CODE EAN
I010120	PINCE LARYNGEE MORS CURETTE D:4MM L:220MM DROITE	3661219401988
	MICRO LARYNGEAL FORCEPS ROUND CUPS D:4MM L:220MM STRAIGHT	 3 661219 401988

INFORMATIONS REGLEMENTAIRES / *REGULATORY INFORMATION*

Classe CE :

I

Code GMDN :

33442

CE Class :

GMDN code :

PERFORMANCES REVENDIQUEES / *INTENDED PERFORMANCE*

Dispositif médical de bloc opératoire :

REUTILISABLE/REUSABLE

Operating room medical device :

CARACTERISTIQUES SPECIFIQUE / *SPECIFIC FEATURES*

Type matière :

ACIER INOXYDABLE / STAINLESS STEEL

Material type :

Nuance matière :

X20Cr13(1.4021) + X5CrNi18-10(1.4301) + X8CrNiS18-9(1.4305) + X10CrNi18-8(1.4310)

Material grade :

Dimensions :

Voir désignations produits

ORL 2021

Outer dimensions :

See products description

Chapitre-pages

319

Nettoyage/Pré-désinfection : Standard pour instrumentation de bloc opératoire réutilisable.

Cleaning / Pre-disinfection : *Standard conditions for reusable surgical instruments*

Stérilisation :

La stérilisation doit être effectuée selon un procédé de stérilisation à la vapeur conformément à la norme ISO 17665-1 (procédé sous vide fractionné) avec un stérilisateur conformément à la norme EN 285.

Méthode de stérilisation à la vapeur saturée	Température	Durée	Temps de séchage	Pression absolue
Pré-vide	132 °C (270 °F)	4 minutes	30 minutes	2,9 bars, 2,8 atm, 41,6 psi
Pré-vide	134 °C (273 °F)	3 minutes	35 minutes	3,0 bars, 3,0 atm, 44,1 psi
Pré-vide	134 °C (273 °F)	5 minutes	35 minutes	3,0 bars, 3,0 atm, 44,1 psi
Pré-vide	134 °C (273 °F)	18 minutes	35 minutes	3,0 bars, 3,0 atm, 44,1 psi

Sterilization process :

The Sterilization shall be performed according to ISO 17665-1 validated steam sterilisation procedure (fractional vacuum procedure) in an autoclave in compliance with EN 285.

Saturated steam sterilization method	Temperature	Time	Drying time	Absolute pressure
Pre-vacuum	132 °C (270 °F)	4 minutes	30 minutes	2.9 bar, 2.8 atm, 41.6 psi
Pre-vacuum	134 °C (273 °F)	3 minutes	35 minutes	3.0 bar, 3.0 atm, 44.1 psi
Pre-vacuum	134 °C (273 °F)	5 minutes	35 minutes	3.0 bar, 3.0 atm, 44.1 psi
Pre-vacuum	134 °C (273 °F)	18 minutes	35 minutes	3.0 bar, 3.0 atm, 44.1 psi

EMBALLAGE / *PACKAGING*

Conditionnement :

Pochette plastique en polyéthylène (PE).

Packaging :

Polyethylene plastic pouch (PE).