

L A N D A N G E R	FICHE TECHNIQUE PRODUIT PRODUCT TECHNICAL SHEET	FAQ 026 INFO Indice : 2 Date : 01/06/2018
-------------------	--	--

Editée le / Publishing date

2022-04-29

Page

1

REFERENCE	DESIGNATIONS - DESCRIPTIONS	CODE EAN
I010151	PINCE LARYNGEE MORS CURETTE D:1MM L:220MM DROITE	3661219402152
	MICRO LARYNGEAL FORCEPS ROUND CUPS D:1MM L:220MM STRAIGHT	 3 661219 402152

Catalogues / catalogs

ORL 2021

Chapitre-page

334

DESCRIPTIF DU DISPOSITIF/ PRODUCT DESCRIPTION

L A N D A N G E R

Micro pinces curettes - Micro curettes forceps - Micro pinzas curetas

170 mm

Ø 1 mm

Droit Straight Recta	Coudé 45° Angled 45° Curva 45°	Courbe droite Right curve Curva derecha	Courbe gauche Left curve Curva izquierda
I010150	I010156	I010152	I010154



I010150
27 cm - 10 1/2"

220 mm

Ø 1 mm

Droit Straight Recta	Coudé 45° Angled 45° Curva 45°	Courbe droite Right curve Curva derecha	Courbe gauche Left curve Curva izquierda
I010151	I010157	I010153	I010155



I010151
32 cm - 12 3/4"

220 mm

Ø 2 mm

Droit Straight Recta	Courbe droite Right curve Curva derecha	Courbe gauche Left curve Curva izquierda
I010110	I010112	I010113
Coudé 45° Angled 45° Curva 45°	Courbe droite Right curve Curva derecha	Courbe gauche Left curve Curva izquierda
I010111	I010114	I010115



I010115
35 cm - 10"

2x 5 mm

I010125



334



1/2

Copyright © LANDANGER - 2018 - 2021

LANDANGER	FICHE TECHNIQUE PRODUIT <i>PRODUCT TECHNICAL SHEET</i>	FAQ 026 INFO Indice : 2 Date : 01/06/2018
-----------	---	---

Éditée le / *Publishing date*

2022-04-29

Page

2

REFERENCE	DESIGNATIONS - <i>DESCRIPTIONS</i>	CODE EAN
I010151	PINCE LARYNGEE MORS CURETTE D:1MM L:220MM DROITE	3661219402152
	MICRO LARYNGEAL FORCEPS ROUND CUPS D:1MM L:220MM STRAIGHT	 3 661219 402152

INFORMATIONS REGLEMENTAIRES / *REGULATORY INFORMATION*

Classe CE :

I

Code GMDN :

35732

CE Class :

GMDN code :

PERFORMANCES REVENDIQUEES / *INTENDED PERFORMANCE*

Dispositif médical de bloc opératoire :

REUTILISABLE/REUSABLE

Operating room medical device :

CARACTERISTIQUES SPECIFIQUE / *SPECIFIC FEATURES*

Type matière :

ACIER INOXYDABLE / STAINLESS STEEL

Material type :

Nuance matière :

X20Cr13(1.4021) + X5CrNi18-10(1.4301) + X8CrNiS18-9(1.4305) + X10CrNi18-8(1.4310)

Material grade :

Dimensions :

Voir désignations produits

ORL 2021

Outer dimensions :

See products description

Chapitre-pages

334

Nettoyage/Pré-désinfection : Standard pour instrumentation de bloc opératoire réutilisable.

Cleaning / Pre-disinfection : *Standard conditions for reusable surgical instruments*

Stérilisation :

La stérilisation doit être effectuée selon un procédé de stérilisation à la vapeur conformément à la norme ISO 17665-1 (procédé sous vide fractionné) avec un stérilisateur conformément à la norme EN 285.

Méthode de stérilisation à la vapeur saturée	Température	Durée	Temps de séchage	Pression absolue
Pré-vide	132 °C (270 °F)	4 minutes	30 minutes	2,9 bars, 2,8 atm, 41,6 psi
Pré-vide	134 °C (273 °F)	3 minutes	35 minutes	3,0 bars, 3,0 atm, 44,1 psi
Pré-vide	134 °C (273 °F)	5 minutes	35 minutes	3,0 bars, 3,0 atm, 44,1 psi
Pré-vide	134 °C (273 °F)	18 minutes	35 minutes	3,0 bars, 3,0 atm, 44,1 psi

Sterilization process :

The Sterilization shall be performed according to ISO 17665-1 validated steam sterilisation procedure (fractional vacuum procedure) in an autoclave in compliance with EN 285.

Saturated steam sterilization method	Temperature	Time	Drying time	Absolute pressure
Pre-vacuum	132 °C (270 °F)	4 minutes	30 minutes	2.9 bar, 2.8 atm, 41.6 psi
Pre-vacuum	134 °C (273 °F)	3 minutes	35 minutes	3.0 bar, 3.0 atm, 44.1 psi
Pre-vacuum	134 °C (273 °F)	5 minutes	35 minutes	3.0 bar, 3.0 atm, 44.1 psi
Pre-vacuum	134 °C (273 °F)	18 minutes	35 minutes	3.0 bar, 3.0 atm, 44.1 psi

EMBALLAGE / *PACKAGING*

Conditionnement :

Pochette plastique en polyéthylène (PE).

Packaging :

Polyethylene plastic pouch (PE).