

L A N D A N G E R	FICHE TECHNIQUE PRODUIT <i>PRODUCT TECHNICAL SHEET</i>	FAQ 026 INFO Indice : 2 Date : 01/06/2018
-------------------	---	---

Éditée le / *Publishing date*

2022-04-29

Page

1

REFERENCE	DESIGNATIONS - <i>DESCRIPTIONS</i>	CODE EAN
I050220	GABARIT COULISSANT OREILLE MOYENNE 0 A 8MM L:180MM	3661219403951
	SLIDABLE MIDDLE EAR GAUGE 0-8MM L:180MM	 3 661219 403951

Catalogues / *catalogs*

ORL 2021

Chapitre-page

177

DESCRIPTIF DU DISPOSITIF/ *PRODUCT DESCRIPTION*

L A N D A N G E R

Micro mesurateurs - Micro measuring rod - Micromonitores

3,50 / 4,00 / 4,50 mm **I050207** 

4,25 / 4,50 / 4,75 mm **I050208**

4,50 / 5,00 / 5,50 mm **I050209**

0-8mm **I050220** 



I050207
16 cm - 6 1/4"

 1/1



I050220
18 cm - 7"



I050235
11,5 cm - 4 1/2"

W - Otolaryngologie
W - Otolaryngologie
W - Otolaryngologie

0,8 0,6
4,75
4,50
4,25
4,00
3,75
3,50
3,25
3,00
2,75

177

LANDANGER	FICHE TECHNIQUE PRODUIT <i>PRODUCT TECHNICAL SHEET</i>	FAQ 026 INFO Indice : 2 Date : 01/06/2018
-----------	---	---

Éditée le / *Publishing date*

2022-04-29

Page

2

REFERENCE	DESIGNATIONS - <i>DESCRIPTIONS</i>	CODE EAN
I050220	GABARIT COULISSANT OREILLE MOYENNE 0 A 8MM L:180MM	3661219403951
	SLIDABLE MIDDLE EAR GAUGE 0-8MM L:180MM	 3 661219 403951

INFORMATIONS REGLEMENTAIRES / *REGULATORY INFORMATION*

Classe CE :

I-MDR

Code GMDN :

46768

CE Class :

GMDN code :

PERFORMANCES REVENDIQUEES / *INTENDED PERFORMANCE*

Dispositif médical de bloc opératoire :

REUTILISABLE/REUSABLE

Operating room medical device :

CARACTERISTIQUES SPECIFIQUE / *SPECIFIC FEATURES*

Type matière :

ACIER INOXYDABLE / STAINLESS STEEL

Material type :

Nuance matière :

X20Cr13(1.4021) + X5CrNi18-10(1.4301) + X8CrNiS18-9(1.4305) + X10CrNi18-8(1.4310)

Material grade :

Dimensions :

Voir désignations produits

ORL 2021

Outer dimensions :

See products description

Chapitre-pages

177

Nettoyage/Pré-désinfection : Standard pour instrumentation de bloc opératoire réutilisable.

Cleaning / Pre-disinfection : *Standard conditions for reusable surgical instruments*

Stérilisation :

La stérilisation doit être effectuée selon un procédé de stérilisation à la vapeur conformément à la norme ISO 17665-1 (procédé sous vide fractionné) avec un stérilisateur conformément à la norme EN 285.

Méthode de stérilisation à la vapeur saturée	Température	Durée	Temps de séchage	Pression absolue
Pré-vide	132 °C (270 °F)	4 minutes	30 minutes	2,9 bars, 2,8 atm, 41,6 psi
Pré-vide	134 °C (273 °F)	3 minutes	35 minutes	3,0 bars, 3,0 atm, 44,1 psi
Pré-vide	134 °C (273 °F)	5 minutes	35 minutes	3,0 bars, 3,0 atm, 44,1 psi
Pré-vide	134 °C (273 °F)	18 minutes	35 minutes	3,0 bars, 3,0 atm, 44,1 psi

Sterilization process :

The Sterilization shall be performed according to ISO 17665-1 validated steam sterilisation procedure (fractional vacuum procedure) in an autoclave in compliance with EN 285.

Saturated steam sterilization method	Temperature	Time	Drying time	Absolute pressure
Pre-vacuum	132 °C (270 °F)	4 minutes	30 minutes	2.9 bar, 2.8 atm, 41.6 psi
Pre-vacuum	134 °C (273 °F)	3 minutes	35 minutes	3.0 bar, 3.0 atm, 44.1 psi
Pre-vacuum	134 °C (273 °F)	5 minutes	35 minutes	3.0 bar, 3.0 atm, 44.1 psi
Pre-vacuum	134 °C (273 °F)	18 minutes	35 minutes	3.0 bar, 3.0 atm, 44.1 psi

EMBALLAGE / *PACKAGING*

Conditionnement :

Pochette plastique en polyéthylène (PE).

Packaging :

Polyethylene plastic pouch (PE).