

L A N D A N G E R	FICHE TECHNIQUE PRODUIT PRODUCT TECHNICAL SHEET	FAQ 026 INFO Indice : 2 Date : 01/06/2018
-------------------	--	--

Editée le / Publishing date

2022-04-29

Page

1

REFERENCE	DESIGNATIONS - <i>DESCRIPTIONS</i>	CODE EAN
J252014C	PINCE EMPORTE PIECE PRESTIGE FERRIS SMITH KERRISON L:200MM MORS 4MM 40° COURBE EN HAUT, DEMONTABLE	3661219428596
	<i>PRESTIGE FERRIS SMITH KERRISON PUNCH FORCEPS L:200MM JAWS 4MM 40° UPWARD CURVED, DETACHABLE</i>	

Catalogues / catalogs

OrthoVertebre 2019

Chapitre-page

279

DESCRIPTIF DU DISPOSITIF/ *PRODUCT DESCRIPTION*

L A N D A N G E R *Prestige*

Pinces emporte pièces - Punches forceps - Pinzas sacabocados

Ferris-Smith-Kerrison
 Démontable - Detachable - Desmontable



40° Haut, Courbe / 40° Upward, Curved / 40° Arriba, Curva

Largueur de mors / Width of jaw / Anchura de boca

1 mm	J252011C
2 mm	J252012C
3 mm	J252013C
4 mm	J252014C

Longueur tige - Shaft length - Largo del tallo

U : 200 mm - 8"

J252012C
 U : 165 mm - 6 1/2"

Copyright © LANDANGER - 2019

U : Longueur utile
 U : Working length
 U : Longitud útil

279

VIII - Pinces
 VIII - Forceps
 VIII - Pinzas

LANDANGER	FICHE TECHNIQUE PRODUIT <i>PRODUCT TECHNICAL SHEET</i>	FAQ 026 INFO Indice : 2 Date : 01/06/2018
-----------	---	---

Éditée le / *Publishing date*

2022-04-29

Page

2

REFERENCE	DESIGNATIONS - <i>DESCRIPTIONS</i>	CODE EAN
J252014C	PINCE EMPORTE PIECE PRESTIGE FERRIS SMITH KERRISON L:200MM MORS 4MM 40° COURBE EN HAUT, DEMONTABLE	3661219428596
	PRESTIGE FERRIS SMITH KERRISON PUNCH FORCEPS L:200MM JAWS 4MM 40° UPWARD CURVED, DETACHABLE	 3 661219 428596

INFORMATIONS REGLEMENTAIRES / *REGULATORY INFORMATION*

Classe CE :
CE Class :

I

Code GMDN :
GMDN code :

32853

PERFORMANCES REVENDIQUEES / *INTENDED PERFORMANCE*

Dispositif médical de bloc opératoire :
Operating room medical device :

REUTILISABLE/REUSABLE

CARACTERISTIQUES SPECIFIQUE / *SPECIFIC FEATURES*

Type matière :
Material type :

ACIER INOXYDABLE / STAINLESS STEEL

Nuance matière :
Material grade :

X20Cr13(1.4021)

Dimensions : Voir désignations produits
Outer dimensions : See products description
Chapitre-pages

OrthoVertebre 2019

279

Nettoyage/Pré-désinfection : Standard pour instrumentation de bloc opératoire réutilisable.
Cleaning / Pre-disinfection : Standard conditions for reusable surgical instruments

Stérilisation :

La stérilisation doit être effectuée selon un procédé de stérilisation à la vapeur conformément à la norme ISO 17665-1 (procédé sous vide fractionné) avec un stérilisateur conformément à la norme EN 285.

Méthode de stérilisation à la vapeur saturée	Température	Durée	Temps de séchage	Pression absolue
Pré-vide	132 °C (270 °F)	4 minutes	30 minutes	2,9 bars, 2,8 atm, 41,6 psi
Pré-vide	134 °C (273 °F)	3 minutes	35 minutes	3,0 bars, 3,0 atm, 44,1 psi
Pré-vide	134 °C (273 °F)	5 minutes	35 minutes	3,0 bars, 3,0 atm, 44,1 psi
Pré-vide	134 °C (273 °F)	18 minutes	35 minutes	3,0 bars, 3,0 atm, 44,1 psi

Sterilization process :

The Sterilization shall be performed according to ISO 17665-1 validated steam sterilisation procedure (fractional vacuum procedure) in an autoclave in compliance with EN 285.

Saturated steam sterilization method	Temperature	Time	Drying time	Absolute pressure
Pre-vacuum	132 °C (270 °F)	4 minutes	30 minutes	2.9 bar, 2.8 atm, 41.6 psi
Pre-vacuum	134 °C (273 °F)	3 minutes	35 minutes	3.0 bar, 3.0 atm, 44.1 psi
Pre-vacuum	134 °C (273 °F)	5 minutes	35 minutes	3.0 bar, 3.0 atm, 44.1 psi
Pre-vacuum	134 °C (273 °F)	18 minutes	35 minutes	3.0 bar, 3.0 atm, 44.1 psi

EMBALLAGE / *PACKAGING*

Conditionnement : Pochette plastique en polyéthylène (PE).
Packaging : Polyethylene plastic pouch (PE).