

SGS Portugal, S.A.
Industrial Services
Pólo Tecnológico de Lisboa
R. Cupertino de Miranda, Lt 6, Piso 0 e 1
1600-546 Lisboa - Portugal
T. (351) 21 710 42 00
F. (351) 21 715 75 26



EC Type-verification

Acc. to Directive 97/23/EC & 2014/68/UE

Certificate n°: **PTC15.02177.5046**

Date: **2015.08.04**

Name and address of the manufacturer: **PROHS - Equipamento Hospitalar e Serviços Associados, S.A.**
Rua do Castanhal N°316, ZI da Maia I Sector II,
4475-122 Maia
Portugal

SGS Portugal SA as a notified body for the scope of Pressure Equipment Directive 97/23/EC & 2014/68/UE, certifies that the Pressure Equipment design, construction and inspection, is in compliance with the requirements of the Pressure Equipment Directive stated, acc. module B.



Assessed acc. to
Directive 97/23/EC & 2014/68/UE:

Module B

Inspection Report n°: **PTR15.04810.5046**

Equipment ID: **Esterilizador Vertical ; 150L ; SN: 3201; Model: EV-150;**

PED data: **PS: -1 / 2.6 bar**
TS: -10 / 138 °C
Category: III
Fluid/Group: Gas - Gr.2
PT: 6.0 bar
Year of Manufacture: 2015
Variants: EV-075; EV-100

Valid Until: **2025.08.04**


José Rabião
Technical Manager



INDIP813/3 (2015-08-03)

This document is issued, on the Client's behalf, by the Company under its General Conditions of Service printed overleaf.
The Client's attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein.

Any other holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Clients instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents.

SGSPAPER
15924561

