

SleepSense®

The Sensor Solutions Company

ThermoCan™ 7-Ft Pediatric Thermal Cannula (Small)

With nasal prongs and integrated thermal sensor.

7' (210 cm) tube with two male luer connectors. **Single use. Do not sterilize.**



Caution, consult Instructions for Use



CE marking of conformity



Do not reuse



Not made with natural rubber latex



BPA Free



DEHP Free



Keep dry



Keep away from sunlight



RoHS Compliance



Use by:

YYYY / MM



00000000

15813*

Item NJ5813 - ver.1.4 (side A)



S.L.P. Ltd.

62 Anilevitz St,
Tel Aviv, 6706016,
ISRAEL
Tel: +972 3 6870640
Fax: +972 3 5371282

S.L.P. Inc.

2511 Technology Dr., Suite 112
Elgin, IL 60124, USA
Tel: +1 (630) 513 7479
Fax: +1 (630) 513 8478
E-mail: info@sleepsense.com

EC REP


Medes Limited

Shenley Hill, Radlett,
Herts WD7 7AR,
England
Tel: +423-663-169205
Tel/Fax: +44 19238 59810

www.sleepsense.com

Toll-free(U.S.): 888-slp-sens (757 7367)

Made in China.



Directions for Use

- Connect the thermal cannula using SleepSense pediatric interface cable by inserting metal plugs on either side of the nose piece. (figure 1)
- Position cannula on face by inserting cannula tips into nostrils. (figure 2)
- Connect one tube to a Pressure Sensor and the other to a Capnograph.
- Connect interface cable to the PSG system.
- Ask the patient to breath through his nose and mouth and ensure that clear signals are recorded.

Indications for use

SleepSense sleep-lab sensors provide a qualitative measure of sounds, respiratory-effort, flow, body position or limb movements, for recording on an approved data acquisition system.

They are intended for use on children and adult patients who are screened during sleep disorder studies at a sleep laboratory or the patient's home.



Warnings and Precautions

- **Caution:** This product is for diagnostic purposes only and is not to be used as an apnea monitor or in a life sustaining situation.
US Federal Law restricts this device to sale, distribution and use by or on the order of a physician.
- SleepSense sensors are for professional use only.
- SleepSense sensors may be used only in conjunction with an approved recording system.
- Discontinue use if sensor shows signs of wear, damage, or exposed metal.
- SleepSense sensors are only intended for use on healthy skin.