



BATCH RELEASE AND CONFORMITY CERTIFICATE
CERTIFICAT DE CONFORMITE ET
DE LIBERATION DE LOT

DIAGAST
251, Av. E. Avinée –
59120 LOOS France
www.diagast.com

REF 69015

LOT

PALERME



CE

SPECIFICITY CONTROL / CONTROLE DE SPECIFICITE

- Technique: According to package insert
- Panel : Red cells of O, A and B groups treated with papain having the following antigens / : Cw, C, c, D, E, e, K, k, Fya, Fyb, Jka, Jkb, Lua, Lub, Kpa, Kpb, Lea, Leb, P1, M, N, S, s.
- SPECIFICATIONS: No agglutination, no hemolysis.

-CONFORMITY YES NO

ACTIVITY CONTROL / CONTROLE DE SENSIBILITE

- Technique : According to package insert
- Panel : Panel of antibodies active in papain 37°C /
- SPECIFICATIONS: Red cells having the antigen which corresponds to the tested antibody
The title obtained with the papain must correspond to the known title (at ± 1 title).

-CONFORMITY YES NO

CONCLUSION

CONFORMITY : YES NO

Sylvie BEGUE

For PRODUCTS RELEASE
QUALITY ASSURANCE SPECIALIST

RELEASE DATE /
DATE LIBERATION :