



Iris

## Certificate of Analysis

### iQ Calibrator

**REF** 800-3103

**LOT** 102-22

**MAN** 2022-04-22

**EXPI** 2022-11-30

Test	Specification	Results
Average Cell Concentration	1200 Cell/ $\mu$ L $\pm$ 100	1178
Specific Gravity @ 20°C	1.007 $\pm$ 0.001	1.007

This lot was manufactured and conforms to the current revision of our specification for this product.

*n. aly*

Quality Control

*26 APR 2022*

Date

