

Certificate of Analysis

Iris Diluent

REF 800-3202**LOT** 265-22 2022-09-29 2023-08-31

Test	Specification	Results
pH	5.7 – 7.5	6.7
Specific Gravity @ 20°C	1.007 ± 0.001	1.008
Clarity	Clear	Clear
Particle Count	≤ 50 / 0.5 mL	6

This lot was manufactured and conforms to the current revision of our specification for this product.

B Rivera

Quality Control

04 OCT 22

Date

