



APC Annexin V Apoptosis Detection Kit with PI

Catalog# / Size 640932 / 100 tests

Regulatory Status RUO

Other Names Annexin A5 Apoptosis Detection Kit

DescriptionBioLegend's APC Annexin V Apoptosis Detection Kit with PI has been specifically designed for the identification

of apoptotic and necrotic cells.

Annexin V (or Annexin A5) is a member of the annexin family of intracellular proteins that binds to phosphatidylserine (PS) in a calcium-dependent manner. PS is normally only found on the intracellular leaflet of the plasma membrane in healthy cells, but during early apoptosis, membrane asymmetry is lost and PS translocates to the external leaflet. Fluorochrome-labeled Annexin V can then be used to specifically target and identify apoptotic cells. Annexin V Binding Buffer is recommended for use with Annexin V staining. Annexin V binding alone cannot differentiate between apoptotic and necrotic cells. To help distinguish between the necrotic and apoptotic cells we recommend use of our Propidium lodide Solution (PI). Early apoptotic cells will exclude PI, while late stage apoptotic cells and necrotic cells will stain positively, due to the passage of these dyes into the nucleus where they bind to DNA.

Propidium iodide is a fluorescent dye that binds to DNA. When excited by 488 nm laser light, it can be detected with in the PE/Texas Red® channel with a bandpass filter 610/10. It is commonly used in evaluation of cell viability or DNA content in cell cycle analysis by flow cytometry.

Product Details

Verified Reactivity All mammalian species

Concentration Lot-specific (to obtain lot-specific concentration and expiration, please enter the lot number in our Certificate of

Analysis online tool.)

Storage & Handling Store between 2°C and 8°C. Do not freeze.

Caution: Propidium lodide Solution is toxigenic and mutagenic; handle with care.

Application FC - Quality tested

Recommended Usage Staining Procedure:

1. Wash cells twice with cold BioLegend's Cell Staining Buffer, and then resuspend cells in Annexin V Binding Buffer at a concentration of 0.25-1.0 x 10⁷ cells/mL.

2. Transfer 100 µL of cell suspension in a 5 mL test tube.

3. Add 5 µL of APC Annexin V.

4. Add 10 μL of Propidium lodide Solution.

5. Gently vortex the cells and incubate for 15 min at room temperature (25°C) in the dark.

 $6. Add 400 \ \mu L$ of Annexin V Binding Buffer to each tube. Analyze by flow cytometry with proper machine settings.

Application Notes Materials Provided:

0.5 ml of APC Annexin V 1 ml of Propidium Iodide Solution 50 ml of Annexin V Binding Buffer

Materials Not Included:

Cell Staining Buffer (Cat. No. 420201)

For a better experience detecting apoptosis, we now recommend <u>Apotracker™</u>. Cell staining with Apotracker™ is Calcium independent. Thus, no special buffers are required, and the protocol can be shortened for single-step co-staining with other reagents.

Application References

(PubMed link indicates BioLegend citation)

- 1. Ranganathan P, et al. 2014. J AM Soc Nephrol. 25:239. PubMed
- 2. Bernhart E, et al. 2014. Neuro Onco. 16:933. PubMed

Product Citations

- 1. Shrestha C,et al. 2017. PLoS One. 10.1371/journal.pone.0186169. PubMed
- 2. Bikorimana JP, et al. 2022. iScience. 25:105537. PubMed
- 3. Zhou L, et al. 2020. Oxid Med Cell Longev. 2685310:2020. PubMed
- 4. Rodrigues M, et al. 2022. Biochem J. 479:1891. PubMed
- 5. Chaturvedi S, et al. 2022. Cell. 185:2086. PubMed
- 6. Baldominos P, et al. 2022. STAR Protoc. 3:101795. PubMed
- 7. Wang H, et al. 2020. Onco Targets Ther. 13:3903. PubMed
- 8. O'Connor CM, et al. 2022. Cancer Res. 82:721. PubMed
- Blackburn CMR, et al. 2021. Atherosclerosis. 330:76. <u>PubMed</u>
 Azad P, et al. 2022. Exp Mol Med. :. <u>PubMed</u>
- 11. Daza–Cajigal V, *et al.* 2019. Front Immunol. 10:2065. <u>PubMed</u>
- 12. Cremasco F, *et al.* 2021. PLoS One. 16:e0241091. <u>PubMed</u>

Antigen Details


Biology Area

Apoptosis/Tumor Suppressors/Cell Death, Cell Biology, Cell Proliferation and Viability, Neuroscience

Gene ID

308

Product Data



Human T leukemia cell line Jurkat, treated (left) or non-treated (right) with BioLegend's antihuman CD95 (EOS9.1) mAb (Cat. No. 305704) for 4 hours at 37°C, then stained with Annexin V- APC for 15 minutes at 37°C in Annexin V Binding buffer. Propidium lodide (PI) (0.03 µg/Test Cat. No. 421301) was added 5 minutes prior to running tubes.

For research use only. Not for diagnostic use. Not for resale. BioLegend will not be held responsible for patent infringement or other violations that may occur with the use of our products.

*These products may be covered by one or more Limited Use Label Licenses (see the BioLegend Catalog or our website, www.biolegend.com/ordering#license). BioLegend products may not be transferred to third parties, resold, modified for resale, or used to manufacture commercial products, reverse engineer functionally similar materials, or to provide a service to third parties without written approval of BioLegend. By use of these products you accept the terms and conditions of all applicable Limited Use Label Licenses. Unless otherwise indicated, these products are for research use only and are not intended for human or animal diagnostic, therapeutic or commercial use.

8999 BioLegend Way, San Diego, CA 92121 www.biolegend.com Toll-Free Phone: 1-877-Bio-Legend (246-5343) Phone: (858) 768-5800 Fax: (877) 455-9587