

ID	KF008	Valid from	01.03.2020	Replaces ID/Rev.	KF008-06 / 01.07.2017	Invalid from		
Revision	07	Author	Dr. Elke Heck	11.12.2019	Approved by/on	Michael Wiehle	18.12.2019	
Certificate of Analysis (CoA)								

Product: **Ascaris lumbricoides IgG**

Prod.No.: **ASCG0020**

Lot: **ASCG-069**

Expiry: **2023-10-31**

Component	Number	Lot	Prod. Date	Expiry Date
MTP	1	MG0020- 069	2022-04	2023-10-31
CONJ	1	KG0020- 069	2022-04	2023-10-31
CONTROL +	1	PG0020- 069	2022-04	2023-10-31
CONTROL -	1	NG0020- 069	2022-04	2023-10-31
CUT OFF	1	CG0020- 069	2022-04	2023-10-31
WASH BUF 20x	1	W- 1034	2022-04	2024-04-30
SOLN STOP	1	S- 964	2022-04	2024-04-30
SUB TMB	1	T- 070	2021-01	2024-01-31
DIL G	1	DG- 466	2022-02	2024-02-29

Component	OD	OD Validation Criteria
CONTROL -	0,001	< 0,200 < Cut-off
CUT OFF	0,609	0,150 – 1,300
CONTROL +	1,241	> Cut-off

Final report issued: 29.04.2022

Approved by: 29.04.2022

Date

Head of Quality Control

