

ID	KF008	Valid from	01.03.2020	Replaces ID/Rev.	KF008-06 / 01.07.2017	Invalid from		
Revision	07	Author	Dr. Elke Heck	11.12.2019	Approved by/on	Michael Wiehle	18.12.2019	
Certificate of Analysis (CoA)								

Product: **Candida albicans IgG**

Prod.No.: **CANG0060**

Lot: **CANG-181**

Expiry: **2024-01-31**

Component	Number	Lot	Prod. Date	Expiry Date
MTP	1	MG0060- 181	2022-07	2024-01-31
CONJ	1	KG0060- 181	2022-07	2024-01-31
CONTROL +	1	PG0060- 181	2022-07	2024-01-31
CONTROL -	1	NG0060- 181	2022-07	2024-01-31
CUT OFF	1	CG0060- 181	2022-07	2024-01-31
WASH BUF 20x	1	W- 1038	2022-05	2024-05-31
SOLN STOP	1	S- 969	2022-05	2024-05-31
SUB TMB	1	T- 070	2021-01	2024-01-31
DIL G	1	DG- 468	2022-05	2024-05-31

Component	OD	OD Validation Criteria
CONTROL -	0,000	< 0,200 < Cut-off
CUT OFF	0,654	0,150 – 1,300
CONTROL +	1,591	> Cut-off

Final report issued: 21.07.2022

Approved by: 21.07.2022

Date

Head of Quality Control

