

ID	KF008	Valid from	01.03.2020	Replaces ID/Rev.	KF008-06 / 01.07.2017	Invalid from		
Revision	07	Author	Dr. Elke Heck	11.12.2019	Approved by/on	Michael Wiehle	18.12.2019	
Certificate of Analysis (CoA)								

Product: **Dengue Virus IgG**

Prod.No.: **DENG0120**

Lot: **DENG-167**

Expiry: **2023-10-31**

Component	Number	Lot	Prod. Date	Expiry Date
MTP	1	MG0120- 167	2022-04	2023-10-31
CONJ	1	KG0120- 167	2022-04	2023-10-31
CONTROL +	1	PG0120- 167	2022-04	2023-10-31
CONTROL -	1	NG0120- 167	2022-04	2023-10-31
CUT OFF	1	CG0120- 167	2022-04	2023-10-31
WASH BUF 20x	1	W- 1035	2022-04	2024-04-30
SOLN STOP	1	S- 965	2022-04	2024-04-30
SUB TMB	1	T- 070	2021-01	2024-01-31
DIL G	1	DG- 467	2022-03	2024-03-31

Component	OD	OD Validation Criteria
CONTROL -	0,004	< 0,200 < Cut-off
CUT OFF	0,620	0,150 – 1,300
CONTROL +	1,827	> Cut-off

Final report issued: 09.05.2022

Approved by: 09.05.2022

Date

Head of Quality Control

