

ID	KF008	Valid from	01.03.2020	Replaces ID/Rev.	KF008-06 / 01.07.2017	Invalid from		
Revision	07	Author	Dr. Elke Heck	11.12.2019	Approved by/on	Michael Wiehle	18.12.2019	
Certificate of Analysis (CoA)								

Product: **Epstein-Barr Virus (EBNA) IgG**

Prod.No.: **EBVG0580**

Lot: **EBVNG-091**

Expiry: **2024-02-29**

Component	Number	Lot	Prod. Date	Expiry Date
MTP	1	MG0580- 091	2022-08	2024-02-29
CONJ	1	KG0580- 091	2022-08	2024-02-29
CONTROL +	1	PG0580- 091	2022-08	2024-02-29
CONTROL -	1	NG0580- 091	2022-08	2024-02-29
CUT OFF	1	CG0580- 091	2022-08	2024-02-29
WASH BUF 20x	1	W- 1038	2022-05	2024-05-31
SOLN STOP	1	S- 969	2022-05	2024-05-31
SUB TMB	1	T- 070-1	2021-01	2025-01-31
DIL G	1	DG- 468	2022-05	2024-05-31

Component	OD	OD Validation Criteria
CONTROL -	0,005	< 0,200 < Cut-off
CUT OFF	0,567	0,150 – 1,300
CONTROL +	1,403	> Cut-off

Final report issued: 10.08.2022

Approved by: 16.08.2022

Date

Head of Quality Control

