

ID	KF008	Valid from	01.03.2020	Replaces ID/Rev.	KF008-06 / 01.07.2017		Invalid from		
Revision	07	Author	Dr. Elke Heck	11.12.2019	Approved by/on	Michael Wiehle	18.12.2019		
Certificate of Analysis (CoA)									

Product: **Entamoeba histolytica IgG**

Prod.No.: **ENTG0140**

Lot: **ENTG-080**

Expiry: **2024-03-31**

Component	Number	Lot	Prod. Date	Expiry Date
MTP	1	MG0140- 080	2022-09	2024-03-31
CONJ	1	KG0140- 080	2022-09	2024-03-31
CONTROL +	1	PG0140- 080	2022-09	2024-03-31
CONTROL -	1	NG0140- 080	2022-09	2024-03-31
CUT OFF	1	CG0140- 080	2022-09	2024-03-31
WASH BUF 20x	1	W- 1042	2022-09	2024-09-30
SOLN STOP	1	S- 973	2022-08	2024-08-31
SUB TMB	1	T- 071	2021-03	2025-03-31
DIL G	1	DG- 469	2022-06	2024-06-30

Component	OD	OD Validation Criteria
CONTROL -	0,015	< 0,200 < Cut-off
CUT OFF	0,530	0,150 – 1,300
CONTROL +	0,951	> Cut-off

Final report issued: 07.10.2022

Approved by: 07.10.2022

Date

Head of Quality Control

