



Certificate of Analysis

Product Name EPO
Product Lot # 3458
Expiration Date 2023-05-31
Storage Temperature 2° to 8°C

CODE #	COMPONENT	LOT#	EXP. DATE
8022	Streptavidin Coated Plates	7923	2023-05
8443	Biotinylated Antibody Solution	8086	2023-06
8444	Enzyme Antibody Solution	8087	2023-06
9707C	Aqueous K-Blue TMB Substrate	8049	2023-06
9624	Wash Buffer (conc.)	8080	2024-06
9753	Stopping Solution	8037	2025-04
8435-8440	EPO Calibrators	7977	2025-01
8441	EPO Control 1	7978	2025-01
8442	EPO Control 2	7979	2025-01

Kit Controls	Vial Ranges (mIU/mL)
Control 1	15.5-25.2
Control 2	144-219

DATA				
CALIBRATOR/CONTROL ID	CONC (mIU/mL)	MEAN OD (450 nm)	MEAN OD (405 nm)	PASS/FAIL
Calibrator A	0	0.011	0.009	Pass
Calibrator B	10.7	0.177	0.064	Pass
Calibrator C	25.9	0.455	0.16	Pass
Calibrator D	53.4	1.029	0.349	Pass
Calibrator E	150	2.744	0.908	Pass
Calibrator F	469	3.961	2.391	Pass
Control 1	19.91	N/A	N/A	Pass
Control 2	205.66	N/A	N/A	Pass

The product listed above has been tested

Product Released By:

N. Fernandez
Quality Control

Date: 24-Jun-22