


TRIDELTA DEVELOPMENT LIMITED**CERTIFICATE OF ANALYSIS**

Product	Purchase order	Lot number	Expiration date
Multispecies SAA		21KT112d	08/23

Lot No.	Description	Standard	Concentration (Bovine Curve)	Average OD
21WCR122	Wash buffer	Std 0	0.00	0.091
21SDB126	Sample diluent	Std 1	18.80	0.207
21SCY091	Anti SAA Conjugate	Std 2	37.50	0.315
21TB318	TMB	Std 3	75.00	0.556
21MP196	Antibody coated plates	Std 4	150.00	0.968
21SA077	Standard	Std 5	300.00	1.827
21ST309	Stop solution			

SAA Calibration curve - SQC**Internal Control:**

Control 1: 112.9 ug/ml (Range 84.1 – 171.4 ug/ml)

Control 2: 76.2 ug/ml (Range 44.8 – 82.5 ug/ml)

Commercial Controls:

Level 1: 41.3 ug/ml (Range 36.3 – 63.9 ng/ml)

Level 2: 125.4 ug/ml (Range 91.6 – 172.4 ng/ml)

Verified by: Paul Mitchell

Experimentation data: 22/09/2022

Accepted

This document has been produced electronically and is valid without a signature