

ID	KF008	Valid from	01.03.2020	Replaces ID/Rev.	KF008-06 / 01.07.2017	Invalid from		
Revision	07	Author	Dr. Elke Heck	11.12.2019	Approved by/on	Michael Wiehle	18.12.2019	
Certificate of Analysis (CoA)								

Product: **Strongyloides**

Prod.No.: **STRO0690**

Lot: **STRO-040**

Expiry: **2023-11-30**

Component	Number	Lot	Prod. Date	Expiry Date
MTP	1	MAB0690- 040	2022-08	2023-11-30
CONJL	1	KAB0690- 040	2022-08	2023-11-30
CONTROL +	1	PAB0690- 040	2022-08	2023-11-30
CONTROL -	1	NAB0690- 040	2022-08	2023-11-30
CUT OFF	1	CAB0690- 040	2022-08	2023-11-30
WASH BUF 20x	1	W- 1039	2022-07	2024-07-31
SOLN STOP	1	S- 970	2022-05	2024-05-31
SUB TMB	1	T- 071	2021-03	2025-03-31
DIL	1	DG- 468	2022-05	2024-05-31

Component	OD	OD Validation Criteria
CONTROL -	0,016	< 0,200 < Cut-off
CUT OFF	0,536	0,150 – 1,300
CONTROL +	1,711	> Cut-off

Final report issued: 25.08.2022

Approved by: 29.08.2022

Date

Head of Quality Control

