

ID	KF008	Valid from	01.03.2020	Replaces ID/Rev.	KF008-06 / 01.07.2017	Invalid from		
Revision	07	Author	Dr. Elke Heck	11.12.2019	Approved by/on	Michael Wiehle	18.12.2019	
Certificate of Analysis (CoA)								

Product: **Trichinella spiralis IgG**

Prod.No.: **TRIG0480**

Lot: **TRIG-079**

Expiry: **2024-02-29**

Component	Number	Lot	Prod. Date	Expiry Date
MTP	1	MG0480- 079	2022-08	2024-02-29
CONJ	1	KG0480- 079	2022-08	2024-02-29
CONTROL +	1	PG0480- 079	2022-08	2024-02-29
CONTROL -	1	NG0480- 079	2022-08	2024-02-29
CUT OFF	1	CG0480- 079	2022-08	2024-02-29
WASH BUF 20x	1	W- 1039	2022-07	2024-07-31
SOLN STOP	1	S- 970	2022-05	2024-05-31
SUB TMB	1	T- 070-1	2021-01	2025-01-31
DIL G	1	DG- 468	2022-05	2024-05-31

Component	OD	OD Validation Criteria
CONTROL -	0,000	< 0,200 < Cut-off
CUT OFF	0,536	0,150 – 1,300
CONTROL +	1,460	> Cut-off

Final report issued: 18.08.2022

Approved by: 18.08.2022

Date

Head of Quality Control



ID	KF008	Valid from	01.03.2020	Replaces ID/Rev.	KF008-06 / 01.07.2017	Invalid from		
Revision	07	Author	Dr. Elke Heck	11.12.2019	Approved by/on	Michael Wiehle	18.12.2019	
Certificate of Analysis (CoA)								

Product: **Trichinella spiralis IgG**

Prod.No.: **TRIG0480**

Lot: **TRIG-079**

Expiry: **2024-02-29**

Component	Number	Lot	Prod. Date	Expiry Date
MTP	1	MG0480- 079	2022-08	2024-02-29
CONJ	1	KG0480- 079	2022-08	2024-02-29
CONTROL +	1	PG0480- 079	2022-08	2024-02-29
CONTROL -	1	NG0480- 079	2022-08	2024-02-29
CUT OFF	1	CG0480- 079	2022-08	2024-02-29
WASH BUF 20x	1	W- 1039	2022-07	2024-07-31
SOLN STOP	1	S- 970	2022-05	2024-05-31
SUB TMB	1	T- 070-1	2021-01	2025-01-31
DIL G	1	DG- 468	2022-05	2024-05-31

Component	OD	OD Validation Criteria
CONTROL -	0,000	< 0,200 < Cut-off
CUT OFF	0,536	0,150 – 1,300
CONTROL +	1,460	> Cut-off

Final report issued: 18.08.2022

Approved by: 18.08.2022

Date

Head of Quality Control

