

## Specification – Certified Reference Material

### Certipur® Certified secondary standard reference buffer solution

pH(S)=6.86<sub>5</sub> (25°C)

#### Accreditation:



Deutsche  
Akkreditierungsstelle  
D-RM-15185-01-00

Merck KGaA, Darmstadt, Germany is accredited by the German accreditation authority as registered reference material producer (D-RM-15185-01-00) in accordance with **ISO 17034**.

<b>Product no.:</b>	1.07202.0105
<b>Description of CRM:</b>	Certified secondary standard reference buffer solution pH(S)=6.86 <sub>5</sub> (25°C) Certified reference material for pH measurement
<b>Expiry date:</b>	3 years
<b>Storage:</b>	+15°C to +25°C tightly closed in the original container
<b>Composition:</b>	Potassium dihydrogen phosphate / di-sodium hydrogen phosphate in water Formulation in compliance with DIN 19266, IUPAC, NIST, Ph. Eur; USP Molality: 0.025 mol/kg / 0.025 mol/kg

Temperature [°C]	Specification as pH	Associated uncertainty $U_{CRM}=k \cdot u_{CRM}$ ( $k=2$ ) as pH
5.0	6.941 – 6.961	± 0.008
10.0	6.913 – 6.933	± 0.008
15.0	6.890 – 6.910	± 0.008
20.0	6.871 – 6.891	± 0.008
<b>25.0</b>	<b>6.855 – 6.875</b>	<b>± 0.008</b>
30.0	6.843 – 6.863	± 0.008
37.0	6.832 – 6.852	± 0.008
40.0	6.828 – 6.848	± 0.009
45.0	6.825 – 6.845	± 0.009
50.0	6.824 – 6.844	± 0.009

<b>Metrological traceability:</b>	This certified secondary standard reference material is directly traceable to primary certified reference material potassium dihydrogen phosphate / di-sodium hydrogen phosphate characterised by PTB-PHOA-xxx/xxxxx/xx and NIST 186 Ix +IIx. <i>PTB: Physikalisch Technische Bundesanstalt, Braunschweig, Germany</i> <i>NIST: National Institute of Standards and Technology, Gaithersburg, USA</i>
-----------------------------------	---



<b>Method of analysis:</b>	The pH value is directly measured by differential potentiometry with the aid of two platinum hydrogen electrodes "quasi without transference" according to IUPAC <sup>1</sup> recommendations against solutions prepared from primary reference materials characterised by PTB and NIST.
<b>Intended use:</b>	This reference material is intended for use as a calibration standard for pH instruments or pH electrodes or as a control sample for measuring the pH value.
<b>Instructions for handling and correct use:</b>	The pH value strongly depends on the temperature. Therefore it is necessary to keep the temperature constant during the measurement. Details concerning the nature of any hazard and appropriate precautions are provided in the material safety data sheet.
<b>Health and safety information:</b>	Please refer to the Safety Data Sheet for detailed information about the nature of any hazard and appropriate precautions to be taken.
<b>Preparation:</b>	This reference material is prepared gravimetrically from potassium dihydrogen phosphate, di-sodium hydrogen phosphate and high purity water. The formulation is compliant to DIN 19266, IUPAC <sup>1</sup> , NIST <sup>2</sup> , Ph. Eur. chapter 2.2.3. and USP chapter <791>.

#### Associated uncertainty:

The expanded uncertainty  $U_{\text{CRM}}$  is calculated as  $U_{\text{CRM}} = k \cdot u_{\text{CRM}}$ , where  $k=2$  is the coverage factor for a 95% coverage probability and  $u_{\text{CRM}}$  is the combined standard uncertainty in accordance to ISO 17034.

The combined uncertainty  $u_{\text{CRM}}$  is derived from combination of the squared uncertainty contributions:

$$u_{\text{CRM}} = \sqrt{u^2_{\text{Characterisation}} + u^2_{\text{Homogeneity}} + u^2_{\text{Stability}}}$$

<b><math>u_{\text{characterisation}}</math>:</b>	is the uncertainty in accordance with DIN EN ISO/IEC 17025 which includes e.g. contributions of the primary reference material and the measuring system.
<b><math>u_{\text{homogeneity}}</math>:</b>	is the between-bottle variation in accordance with ISO 17034. The assessment of homogeneity is performed by analysis of a representative number of systematically chosen sample units.
<b><math>u_{\text{stability}}</math>:</b>	is the uncertainty obtained from short-term and long-term stability in accordance with ISO 17034. The stability studies are the basis for the quantification of the expiry date of this reference material for the unopened bottle.

**Detailed information is provided by the certificates and the certification report on our website.**

<sup>1</sup> R.P. Buck, et al.: The Measurement of pH – **Definition, Standards and Procedures (IUPAC Recommendations 2002)**, Pure Appl. Chem, Vol 74, No. 11, pp. 2169-2200, 2002

<sup>2</sup> Y. Ch. Wu, W. F. Koch, R. A. Durst: **Standardization of pH Measurements**, NBS Special Publication 260-53, 1988

