

DECLARATION OF CONFORMITY

We

Manufacturer Canon Inc.
9-1, Imaikami-cho, Nakahara-ku, Kawasaki, Kanagawa 211-8501, Japan

Authorized representative Canon Europa N.V. Medical Imaging Group
in Europe Bovenkerkerweg 59, 1185 XB Amstelveen, The Netherlands

declare under our sole responsibility that product

DIGITAL RADIOGRAPHY model: CXDI-710C Wireless
(Class IIa, Rule10)

is in conformity with essential requirements of EC Directive(s)

93/42/EEC

by applying the following harmonized standards

EC Directive(s)	Reference of standard(s) and amendment(s)
93/42/EEC	EN ISO13485:2016/ AC:2018
	EN 60601-1:2006/ A11:2011/ A1:2013
	EN 60601-1-2:2015
	EN 60601-1-3:2008/A11:2016
	EN 60601-1-6:2010
	EN 60601-2-54:2009
	EN 62220-1:2004
	EN 62304:2006/AC:2008
	EN 62366:2008
	EN ISO 14971:2012
	EN ISO 10993-1:2009/AC:2010, EN ISO 10993-5:2009,
	EN ISO 10993-10:2010, EN ISO 10993-12:2012, EN ISO 10993-18:2009
	EN ISO 15223-1:2016, EN 1041:2008

and a "Full Quality Assurance System" according to EC Directive 93/42/EEC Annex II except Section 4 by the following Notified Body as stated on the certificate of HD 1230912-1.


Notified Body:

Name TUV Rheinland LGA Products GmbH (ID No. 0197)
Address Tillystraße 2 90431 Nürnberg Germany

Note:

- The CE Marking is affixed from the year '17.
- This statement of conformity is only valid in conjunction with the release document for each serial number of produced devices.
- The contents marked with an asterisk is that added or revised to our previous Declaration of Conformity No. 17-M01-11.
- Copies of the rating plates of these models are attached.
- Accessory: (Option)
CXDI Control Software NE, Battery Pack LB-4A, Multi Box MB-4A, Docking Station DS-4A,
Ready Indicator RI-3A, PC Connection Cable CP-4A, Status Indicator SI-4A, Wiring Cable WC-4A, Scatter Correction for CXDI Series, One Shot LL for CXDI Series, Advanced Edge Enhancement, Free Rotation for CXDI Series, *Intelligent NR

Date: November26, 2021



Akira HIRAI
Medical Equipment Quality Assurance Div.
Medical Components Group
CANON INC.

RECORD OF DECLARATION OF CONFORMITY FOR MEDICAL DEVICE

Date	Contents of Technical Report		QA System ISO 13485	Essential Requirement	Comments
	Electrical Safety	EMC			
2017.02.23	Q40-1129 50050814 001	11355757S-A	HD 60113491 0001	Feb 27, 2017	<u>Suffix No. 01</u> is issued. (Model CXDI-710C Wireless new registration)
2017.04.28	-	-	-	April 28, 2017	<u>Suffix No. 02</u> is issued. (Deletion of battery charger BC-1A)
2017.12.25	-	-	-	December. 25, 2017	<u>Suffix No. 03</u> is issued. - Update of the harmonized standard
2018.9.25	-	-	-	September 21, 2018	<u>Suffix No. 04</u> is issued. - Changing Manufacturer name
2019.1.23	-	-	-	January 23, 2019	<u>Suffix No. 05</u> is issued. - Adding option Advanced Edge Enhancement
2019.4.3	—	—	—	April 3, 2019	<u>Suffix No. 06</u> is issued. - Update of the harmonized standard (applicable to EN ISO13485:2016)
2019.7.18	—	—	HD 60140845 0001	—	<u>Suffix No. 07</u> is issued. - Update of the EC Certificate
2019.8.6	—	—	—	August 6, 2019	<u>Suffix No. 08</u> is issued. - Update of the harmonized standards
2019.12.20	—	—	—	—	<u>Suffix No. 09</u> is issued. - Update of the optional function of software
2020.10.26	—	—	—	October 26, 2020	<u>Suffix No. 10</u> is issued. - Update of the harmonized standard
2020.11.23	—	—	HD 1230912-1	—	<u>Suffix No. 11</u> is issued. - Update of full quality assurance system certificate
2021.11.26	—	—	—	November 18, 2021	<u>Suffix No. 12</u> is issued. - Adding “Intelligent NR”