

DECLARATION OF CONFORMITY



Manufacturer

Fort Henry Business Park, Techno-path Manufacturing Ltd.

Ballina,

Co. Tipperary,

Ireland

Product Name

TECHNOPATH Multi-CHECK Combi Level 2 TECHNOPATH Multi-CHECK Combi Level 1

Category

Catalogue Number

Assayed//Single level

Assayed/Single level

CC911A CC912A

GMDN:

47869

Quality Management System: Conformity Route: Annex III Self-Declared

EN ISO 13485:2016

TÜV SÜD, Ridlerstraße 65, 80339 Munich, Germany Q51038520004 Rev 01

QMS Certification No.:

Expiry Date: Issued By:

12 February 2025

compliance can be provided Standards Applied: See attached list of standards for which documented evidence of

with the requirements listed in European Union In-vitro Diagnostic Medical Device Directive 98/79/EC. Techno-path Manufacturing Ltd. hereby declares that the product(s) specified above comply

I am fully responsible for all the information provided in this declaration. This declaration of conformity is valid from 18 (Day) 作品 (Month) 622 (Year)

Signed for and on behalf of Techno-path Manufacturing Ltd.,

Bernd Hass

Place and Date of Issue Ballina, Co.Tipperary

SVP of Quality and Regulatory Affairs

Techno-path Manufacturing Ltd.

Rev 02

DC-EU



STANDARDS USED IN FULL OR PART FOR CE MARKING AS PER IVDD 98/79/EC.

of in vitro diagnostic reagents	
In vitro diagnostic medical devices - Evaluation of stability	EN ISO 23640:2015
reagents for professional use	
by the manufacturer (labelling) – Part 2: In vitro diagnostic	
In vitro diagnostic medical devices – Information supplied	EN ISO 18113-2:2011
and general requirements	
by the manufacturer (labelling) – Part 1: Terms, definitions	
In vitro diagnostic medical devices – Information supplied	EN ISO 18113-1:2011
medical devices	
Medical devices – Application of risk management to	ISO 14971:2019
vitro diagnostic medical devices – statistical aspects	=
Sampling procedures used for acceptance testing of in	EN 13975:2003
vitro diagnostic reagents	
Elimination or reduction of risk of infection related to in	EN 13641:2002
devices	
Performance evaluation of in vitro diagnostic medical	EN 13612:2002 + AC:2002
Requirements for regulatory purposes	
Medical devices – Quality management systems –	EN ISO13485:2016
and information to be supplied.	
Symbols to be used with medical device labels, labelling	EN ISO15223-1:2016
Title	Standard